



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112036

Ambient Barometric Pressure (mmHg): 771.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	79.9	11/18/2020	09:35
2	L4766-12		1	8	100.00	>212.0	>212.0	11/18/2020	10:05
3	L4766-12DUP		1	8	100.00	>212.0	>212.0	11/18/2020	10:40
4	L4767-02		1	9	100.00	>212.0	>212.0	11/18/2020	11:10
5	L4769-11		1	8	100.00	>212.0	>212.0	11/18/2020	11:45
6	L4775-01		1	9	100.00	>212.0	>212.0	11/18/2020	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP111620 WorkList ID : 144199 Department : Wet-Chemistry Date : 11-16-2020 17:02:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/19/2020	Water	L4766-12	Flash Point	Cool 4 deg C	PSEG03	H11	TOTE-COMP-AUD	11/16/2020	1010A
11/19/2020	Water	L4767-02	Flash Point	Cool 4 deg C	PSEG03	H11	NWB-1555	11/16/2020	1010A
11/30/2020	Water	L4769-11	Flash Point	Cool 4 deg C	PSEG03	H11	TRE-1119	11/16/2020	1010A
11/24/2020	Water	L4775-01	Flash Point	Cool 4 deg C	PSEG03	H12	FMI-124	11/17/2020	1010A

Date/Time 11/18/20 09:05
 Received by: KS (Wetchem)
 Relinquished by: SP (Wetchem)

Date/Time 11/18/20 12:25
 Received by: SP (Wetchem)
 Relinquished by: KS (Wetchem)