



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112037

Ambient Barometric Pressure (mmHg): 771.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	79.9	11/18/2020	13:05
2	L4765-03		1	8	100.00	>212.0	>212.0	11/18/2020	13:35
3	L4765-03DUP		1	8	100.00	>212.0	>212.0	11/18/2020	14:10
4	L4778-01		1	8	55.00	131.0	130.3	11/18/2020	14:40
5	L4778-04		1	9	100.00	>212.0	>212.0	11/18/2020	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB112037

WorkList Name : fp111820 WorkList ID : 144262 Department : Wet-Chemistry Date : 11-18-2020 09:17:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2020	Solid	L4765-03	Flash Point	Cool 4 deg C	PSEG03	H11	CF-OIL	11/16/2020	1010A
11/24/2020	Solid	L4778-01	Flash Point	Cool 4 deg C	PSEG03	H12	1840	11/17/2020	1010A
11/24/2020	Solid	L4778-04	Flash Point	Cool 4 deg C	PSEG03	H12	1811	11/17/2020	1010A

Date/Time 11/18/20 09:25
 Received by: KS (watches)
 Relinquished by: 78 (watches) Chem

Date/Time 11/18/20 15:20
 Received by: 20 (watches) Chem
 Relinquished by: KS (watches)