

7470A


LB112043

INSTRUMENT ID: CV2

Linear

 μ Abs.:

36022



A= 0.0000e+000

B= 2.7437e-004

C= 5.5216e-002

Rho= 0.9998262

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	-82	0.000	-82				
0.2	0.200	0.216	0.016	587	0.0 %	587				
2.5	2.500	2.384	-0.116	8487	0.0 %	8487				
5.0	5.000	5.037	0.037	18156	0.0 %	18156				
7.5	7.500	7.592	0.092	27469	0.0 %	27469				
10.0	10.000	9.939	-0.061	36022	0.0 %	36022				

LB112043 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Nov 2020 09:36:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-82	-	0.0000	7470A	PPB	19 Nov 2020 09:38:49	S	Std
0.2 - 1	50.2	587	-	0.2000	7470A	PPB	19 Nov 2020 09:41:06	S	Std
2.5 - 1	52.5	8487	-	2.5000	7470A	PPB	19 Nov 2020 09:43:24	S	Std
5.0 - 1	55	18156	-	5.0000	7470A	PPB	19 Nov 2020 09:45:41	S	Std
7.5 - 1	57.5	27469	-	7.5000	7470A	PPB	19 Nov 2020 09:47:58	S	Std
10.0 - 1	510	36022	-	10.0000	7470A	PPB	19 Nov 2020 09:50:13	S	Std
ICV11 - 1	ICV11	13626	3.7938	-	7470A	PPB	19 Nov 2020 09:52:55	U	SMPL
ICB11 - 1	ICB11	-126	0.0206	-	7470A	PPB	19 Nov 2020 09:55:10	U	SMPL
CCV55 - 1	CCV55	18289	5.0732	-	7470A	PPB	19 Nov 2020 09:57:25	U	SMPL
CCB55 - 1	CCB55	-119	0.0226	-	7470A	PPB	19 Nov 2020 09:59:40	U	SMPL
CRA - 1	CRA	633	0.2289	-	7470A	PPB	19 Nov 2020 10:01:55	U	SMPL
HighStd - 1	HighStd	36686	10.1208	-	7470A	PPB	19 Nov 2020 10:04:10	U	SMPL
ChkStd - 1	ChkStd	25827	7.1414	-	7470A	PPB	19 Nov 2020 10:06:25	U	SMPL
PB133094BL - 1	PBW	-158	0.0119	-	7470A	PPB	19 Nov 2020 10:08:42	U	SMPL
PB133094BS - 1	LCSW	14184	3.9469	-	7470A	PPB	19 Nov 2020 10:13:34	U	SMPL
L4776-02 - 1	ND-SOIL	-102	0.0272	-	7470A	PPB	19 Nov 2020 10:15:49	U	SMPL
L4776-02DUP - 1	ND-SOILDUP	-25	0.0484	-	7470A	PPB	19 Nov 2020 10:18:06	U	SMPL
L4776-02MS - 1	ND-SOILMS	16291	4.5250	-	7470A	PPB	19 Nov 2020 10:20:23	U	SMPL
L4776-02MSD - 1	ND-SOILMSD	15291	4.2506	-	7470A	PPB	19 Nov 2020 10:22:40	U	SMPL
L4778-04 - 1	1811	10	0.0580	-	7470A	PPB	19 Nov 2020 10:24:56	U	SMPL
CCV56 - 1	CCV56	18163	5.0386	-	7470A	PPB	19 Nov 2020 10:27:11	U	SMPL
CCB56 - 1	CCB56	-100	0.0278	-	7470A	PPB	19 Nov 2020 10:29:26	U	SMPL
L4778-06 - 1	VNJ-223	-5	0.0538	-	7470A	PPB	19 Nov 2020 10:31:41	U	SMPL
L4778-07 - 1	358	15	0.0593	-	7470A	PPB	19 Nov 2020 10:33:56	U	SMPL
PB133095BL - 1	PBW	28	0.0629	-	7470A	PPB	19 Nov 2020 10:36:12	U	SMPL
PB133095BS - 1	LCSW	14328	3.9864	-	7470A	PPB	19 Nov 2020 10:38:28	U	SMPL
L4778-08 - 1	1841	202	0.1106	-	7470A	PPB	19 Nov 2020 10:40:43	U	SMPL
L4778-11 - 1	1814	-85	0.0319	-	7470A	PPB	19 Nov 2020 10:43:00	U	SMPL
L4778-11DUP - 1	1814DUP	-9	0.0527	-	7470A	PPB	19 Nov 2020 10:45:16	U	SMPL
L4778-11MS - 1	1814MS	12714	3.5436	-	7470A	PPB	19 Nov 2020 10:47:33	U	SMPL
L4778-11MSD - 1	1814MSD	13940	3.8799	-	7470A	PPB	19 Nov 2020 10:49:50	U	SMPL
L4779-01 - 1	1643	-108	0.0256	-	7470A	PPB	19 Nov 2020 10:52:06	U	SMPL
CCV57 - 1	CCV57	17961	4.9832	-	7470A	PPB	19 Nov 2020 10:54:23	U	SMPL
CCB57 - 1	CCB57	-81	0.0330	-	7470A	PPB	19 Nov 2020 10:56:39	U	SMPL
L4779-02 - 1	1641	58	0.0711	-	7470A	PPB	19 Nov 2020 10:58:55	U	SMPL
L4779-03 - 1	1623	9	0.0577	-	7470A	PPB	19 Nov 2020 11:01:10	U	SMPL
L4779-04 - 1	1608	0	0.0552	-	7470A	PPB	19 Nov 2020 11:03:26	U	SMPL
L4779-05 - 1	1555	14	0.0591	-	7470A	PPB	19 Nov 2020 11:05:43	U	SMPL
L4779-13 - 1	COMP-1654-1658	5	0.0566	-	7470A	PPB	19 Nov 2020 11:08:00	U	SMPL
L4779-14 - 1	COMP-1605-1607	-3	0.0544	-	7470A	PPB	19 Nov 2020 11:10:17	U	SMPL
PB133096BL - 1	PBW	-30	0.0470	-	7470A	PPB	19 Nov 2020 11:12:34	U	SMPL
PB133096BS - 1	LCSW	13335	3.7140	-	7470A	PPB	19 Nov 2020 11:14:51	U	SMPL
L4775-01 - 1	FMI-124	9	0.0577	-	7470A	PPB	19 Nov 2020 11:17:08	U	SMPL
L4775-01DUP - 1	FMI-124DUP	-83	0.0324	-	7470A	PPB	19 Nov 2020 11:19:24	U	SMPL
CCV58 - 1	CCV58	18534	5.1404	-	7470A	PPB	19 Nov 2020 11:31:46	U	SMPL
CCB58 - 1	CCB58	-53	0.0407	-	7470A	PPB	19 Nov 2020 11:34:01	U	SMPL
L4775-01MS - 1	FMI-124MS	14720	4.0940	-	7470A	PPB	19 Nov 2020 11:36:16	U	SMPL
L4775-01MSD - 1	FMI-124MSD	14622	4.0671	-	7470A	PPB	19 Nov 2020 11:38:32	U	SMPL
PB133094TB - 1	PB133071TB	-84	0.0322	-	7470A	PPB	19 Nov 2020 11:40:48	U	SMPL
PB133095TB - 1	PB133089TB	-3	0.0544	-	7470A	PPB	19 Nov 2020 11:43:04	U	SMPL
L4776-02LX5 - 1		-2	0.0547	-	7470A	PPB	19 Nov 2020 11:45:20	U	SMPL
L4776-02A - 1		16079	4.4668	-	7470A	PPB	19 Nov 2020 11:47:36	U	SMPL
L4778-11LX5 - 1		-76	0.0344	-	7470A	PPB	19 Nov 2020 11:49:53	U	SMPL
L4778-11A - 1		14420	4.0116	-	7470A	PPB	19 Nov 2020 11:52:09	U	SMPL
L4775-01LX5 - 1		-77	0.0341	-	7470A	PPB	19 Nov 2020 11:54:26	U	SMPL
L4775-01A - 1		14517	4.0383	-	7470A	PPB	19 Nov 2020 11:56:43	U	SMPL
CCV59 - 1	CCV59	19158	5.3116	-	7470A	PPB	19 Nov 2020 11:58:58	U	SMPL
CCB59 - 1	CCB59	-89	0.0308	-	7470A	PPB	19 Nov 2020 12:01:13	U	SMPL