



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112047

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/19/2020	12:05
2	CAL2	1	Water	NA	NA	20.1	7.00	11/19/2020	12:06
3	CAL3	1	Water	NA	NA	20.1	10.03	11/19/2020	12:07
4	ICV	1	Water	NA	NA	20.1	6.98	11/19/2020	12:08
5	CCV1	1	Water	NA	NA	20.2	2.01	11/19/2020	12:10
6	L4820-02	1	Water	NA	NA	15.4	9.41	11/19/2020	12:11
7	L4820-02DUP	1	Water	NA	NA	15.5	9.42	11/19/2020	12:12
8	L4820-03	1	Water	NA	NA	16.9	7.45	11/19/2020	12:15
9	L4820-04	1	Water	NA	NA	18.1	9.14	11/19/2020	12:16
10	CCV2	1	Water	NA	NA	20.1	12.02	11/19/2020	12:17

WORKLIST(Hardcopy Internal Chain)

LB 112647

WorkList Name : PH L4820

WorkList ID : 144302

Department : Wet-Chemistry

Date : 11-19-2020 11:51:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/18/2020	Water	L4820-02	pH	Cool 4 deg C	TETR06	H23	TT022-IDWGW-20201118	11/18/2020	9040C
11/18/2020	Water	L4820-03	pH	Cool 4 deg C	TETR06	H23	TT-023-IDWGW-20201118	11/18/2020	9040C
11/18/2020	Water	L4820-04	pH	Cool 4 deg C	TETR06	H23	TT-024-IDWGW-20201118	11/18/2020	9040C

Date/Time 11-19-2020 12:00
 Received by: JP (wet chem)
 Relinquished by: CL (SM)

Date/Time 11-19-2020 13:30
 Received by: CL (SM)
 Relinquished by: JP (wet chem)