



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112050

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/19/2020	13:25
2	CAL2	1	Water	NA	NA	20.3	7.00	11/19/2020	13:26
3	CAL3	1	Water	NA	NA	20.3	10.02	11/19/2020	13:27
4	ICV	1	Water	NA	NA	20.4	6.99	11/19/2020	13:30
5	CCV1	1	Water	NA	NA	20.4	2.01	11/19/2020	13:32
6	L4826-02	1	Water	NA	NA	13.4	11.43	11/19/2020	13:35
7	L4826-02DUP	1	Water	NA	NA	13.5	11.44	11/19/2020	13:36
8	CCV2	1	Water	NA	NA	20.3	12.02	11/19/2020	13:28

11/20/20

WORKLIST(Hardcopy Internal Chain)

Date : 11-19-2020 13:13:53

Department : Wet-Chemistry

WorkList ID : 144313

WorkList Name : corrosivity 14826 W

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/18/2020	Water	L4826-02	-F Corrosivity	Cool 4 deg C	ARCA05	H33	PB-2020-WW-000-04	11/18/2020	9040C

Date/Time 11-19-2020 13:13:53
Received by: JH (alt - Chem)
Relinquished by: JH

Date/Time 11-19-2020 14:20
Received by: JH
Relinquished by: JH (alt - Chem)