

LB112051


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5511E



A= 0.0000e+000

B= 1.8033e-004

C= 1.3258e-002

Rho= 0.9998648

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	45	0.000	45				
0.2	0.200	0.216	0.016	1123	0.0 %	1123				
2.5	2.500	2.457	-0.043	13553	0.0 %	13553				
5.0	5.000	4.937	-0.063	27303	0.0 %	27303				
7.5	7.500	7.616	0.116	42157	0.0 %	42157				
10.0	10.000	9.953	-0.047	55118	0.0 %	55118				

LB112051 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 19 Nov 2020 13:37:41					
Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.00007471B	PPB		19 Nov 2020 13:38:58	S	Std
0.2 - 1	50.2	1123	-	0.20007471B	PPB		19 Nov 2020 13:41:15	S	Std
2.5 - 1	52.5	13553	-	2.50007471B	PPB		19 Nov 2020 13:43:32	S	Std
5.0 - 1	55	27303	-	5.00007471B	PPB		19 Nov 2020 13:45:49	S	Std
7.5 - 1	57.5	42157	-	7.50007471B	PPB		19 Nov 2020 13:48:05	S	Std
10.0 - 1	510	55118	-	10.00007471B	PPB		19 Nov 2020 13:50:22	S	Std
ICV12 - 1	ICV12	20880	3.7786	-7471B	PPB		19 Nov 2020 13:53:03	U	SMPL
ICB12 - 1	ICB12	-57	0.0030	-7471B	PPB		19 Nov 2020 13:55:18	U	SMPL
CCV60 - 1	CCV60	27141	4.9077	-7471B	PPB		19 Nov 2020 13:57:33	U	SMPL
CCB60 - 1	CCB60	-82	-0.0015	-7471B	PPB		19 Nov 2020 13:59:48	U	SMPL
CRA - 1	CRA	1071	0.2064	-7471B	PPB		19 Nov 2020 14:02:04	U	SMPL
HighStd - 1	HighStd	55068	9.9439	-7471B	PPB		19 Nov 2020 14:04:19	U	SMPL
ChkStd - 1	ChkStd	38460	6.9489	-7471B	PPB		19 Nov 2020 14:06:34	U	SMPL
PB133107BL - 1	PBS	-114	-0.0073	-7471B	PPB		19 Nov 2020 14:08:51	U	SMPL
PB133107BS - 1	LCSS	20500	3.7101	-7471B	PPB		19 Nov 2020 14:11:09	U	SMPL
L4716-02 - 1	STOCKPILE-2	10960	1.9897	-7471B	PPB		19 Nov 2020 14:13:25	U	SMPL
L4786-01 - 1	KL-SOIL	11780	2.1376	-7471B	PPB		19 Nov 2020 14:15:41	U	SMPL
L4797-01 - 1	R-LA-2-5-WC	725	0.1440	-7471B	PPB		19 Nov 2020 14:17:58	U	SMPL
L4800-03 - 1	KENS-B1-111820-0-1	2623	0.4863	-7471B	PPB		19 Nov 2020 14:20:15	U	SMPL
L4800-05 - 1	KENS-B1-111820-8-10	492	0.1020	-7471B	PPB		19 Nov 2020 14:22:30	U	SMPL
CCV61 - 1	CCV61	26515	4.7948	-7471B	PPB		19 Nov 2020 14:24:45	U	SMPL
CCB61 - 1	CCB61	-62	0.0021	-7471B	PPB		19 Nov 2020 14:27:01	U	SMPL
L4802-01 - 1	FK-GF-02-064-A	6210	1.1331	-7471B	PPB		19 Nov 2020 14:29:16	U	SMPL
L4802-03 - 1	FK-GF-02-064-B	6724	1.2258	-7471B	PPB		19 Nov 2020 14:31:31	U	SMPL
L4803-01 - 1	FK-BPM-01-C5-D5-C5A-D5A	848	0.1662	-7471B	PPB		19 Nov 2020 14:33:46	U	SMPL
L4804-01 - 1	FK-BPM-04-012-ABCD	947	0.1840	-7471B	PPB		19 Nov 2020 14:36:02	U	SMPL
L4805-01 - 1	FK-BPM-03-017-ABCD	1477	0.2796	-7471B	PPB		19 Nov 2020 14:38:18	U	SMPL
L4806-01 - 1	FK-BPM-03-018-ABCD	1444	0.2737	-7471B	PPB		19 Nov 2020 14:40:34	U	SMPL
L4806-01DUP - 1	FK-BPM-03-018-ABCDUP	1621	0.3056	-7471B	PPB		19 Nov 2020 14:42:50	U	SMPL
L4806-01MS - 1	FK-BPM-03-018-ABCDMS	23657	4.2794	-7471B	PPB		19 Nov 2020 14:45:07	U	SMPL
L4806-01MSD - 1	FK-BPM-03-018-ABCDMSD	20763	3.7575	-7471B	PPB		19 Nov 2020 14:47:25	U	SMPL
CCV62 - 1	CCV62	26854	4.8560	-7471B	PPB		19 Nov 2020 14:49:40	U	SMPL
CCB62 - 1	CCB62	-69	0.0008	-7471B	PPB		19 Nov 2020 14:51:56	U	SMPL
L4806-01LX5 - 1		262	0.0605	-7471B	PPB		19 Nov 2020 14:54:11	U	SMPL
L4806-01A - 1		23465	4.2448	-7471B	PPB		19 Nov 2020 14:56:27	U	SMPL
CCV63 - 1	CCV63	27066	4.8942	-7471B	PPB		19 Nov 2020 14:58:42	U	SMPL
CCB63 - 1	CCB63	-58	0.0028	-7471B	PPB		19 Nov 2020 15:00:58	U	SMPL