

LB112064


INSTRUMENT ID: CV2

7470A

Linear

 μ Abs.:

36994



A= 0.0000e+000

B= 2.7275e-004

C= -7.8918e-003

Rho= 0.9997295

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	23	0.000	23				
0.2	0.200	0.189	-0.011	723	0.0 %	723				
2.5	2.500	2.530	0.030	9306	0.0 %	9306				
5.0	5.000	5.075	0.075	18637	0.0 %	18637				
7.5	7.500	7.325	-0.175	26884	0.0 %	26884				
10.0	10.000	10.082	0.082	36994	0.0 %	36994				

LB112064 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Nov 2020 10:07:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7470A	PPB	20 Nov 2020 10:35:36	S	Std
0.2 - 1	50.2	723	-	0.2000	7470A	PPB	20 Nov 2020 10:37:53	S	Std
2.5 - 1	52.5	9306	-	2.5000	7470A	PPB	20 Nov 2020 10:40:11	S	Std
5.0 - 1	55	18637	-	5.0000	7470A	PPB	20 Nov 2020 10:42:29	S	Std
7.5 - 1	57.5	26884	-	7.5000	7470A	PPB	20 Nov 2020 10:44:45	S	Std
10.0 - 1	59.0	36994	-	10.0000	7470A	PPB	20 Nov 2020 10:47:01	S	Std
ICV13 - 1	ICV13	13891	3.7808	-	7470A	PPB	20 Nov 2020 10:49:46	U	SMPL
ICB13 - 1	ICB13	-74	-0.0281	-	7470A	PPB	20 Nov 2020 10:52:01	U	SMPL
CCV64 - 1	CCV64	18427	5.0180	-	7470A	PPB	20 Nov 2020 10:56:49	U	SMPL
CCB64 - 1	CCB64	-91	-0.0327	-	7470A	PPB	20 Nov 2020 10:59:04	U	SMPL
CRA - 1	CRA	702	0.1836	-	7470A	PPB	20 Nov 2020 11:01:19	U	SMPL
HighStd - 1	HighStd	36220	9.8710	-	7470A	PPB	20 Nov 2020 11:03:34	U	SMPL
ChkStd - 1	ChkStd	26072	7.1032	-	7470A	PPB	20 Nov 2020 11:05:49	U	SMPL
PB133116BL - 1	PBW	-121	-0.0409	-	7470A	PPB	20 Nov 2020 11:08:06	U	SMPL
PB133116BS - 1	LCSW	14344	3.9044	-	7470A	PPB	20 Nov 2020 11:10:22	U	SMPL
L4786-02 - 1	KL-SOIL	-61	-0.0245	-	7470A	PPB	20 Nov 2020 11:12:39	U	SMPL
L4786-02DUP - 1	KL-SOIL DUP	-6	-0.0095	-	7470A	PPB	20 Nov 2020 11:14:55	U	SMPL
L4786-02MS - 1	KL-SOIL MS	13901	3.7836	-	7470A	PPB	20 Nov 2020 11:17:12	U	SMPL
L4786-02MSD - 1	KL-SOIL MSD	12911	3.5135	-	7470A	PPB	20 Nov 2020 11:19:29	U	SMPL
PB133120BL - 1	PBW	-71	-0.0273	-	7470A	PPB	20 Nov 2020 11:21:44	U	SMPL
CCV65 - 1	CCV65	18220	4.9616	-	7470A	PPB	20 Nov 2020 11:24:00	U	SMPL
CCB65 - 1	CCB65	-66	-0.0259	-	7470A	PPB	20 Nov 2020 11:26:15	U	SMPL
PB133120BS - 1	LCSW	12801	3.4835	-	7470A	PPB	20 Nov 2020 11:28:30	U	SMPL
L4800-06 - 1	FIELD-BLANK	-71	-0.0273	-	7470A	PPB	20 Nov 2020 11:30:45	U	SMPL
L4820-02 - 1	TT022-IDWGW-20201118	30	0.0003	-	7470A	PPB	20 Nov 2020 11:33:00	U	SMPL
L4820-02DUP - 1	TT022-IDWGW-20201118DUP	54	0.0068	-	7470A	PPB	20 Nov 2020 11:35:16	U	SMPL
L4820-02MS - 1	TT022-IDWGW-20201118MS	16546	4.5050	-	7470A	PPB	20 Nov 2020 11:37:32	U	SMPL
L4820-02MSD - 1	TT022-IDWGW-20201118MSD	15068	4.1019	-	7470A	PPB	20 Nov 2020 11:39:49	U	SMPL
L4820-03 - 1	TT-023-IDWGW-20201118	48	0.0052	-	7470A	PPB	20 Nov 2020 11:42:05	U	SMPL
L4820-04 - 1	TT-024-IDWGW-20201118	162	0.0363	-	7470A	PPB	20 Nov 2020 11:44:22	U	SMPL
L4826-02 - 1	PB-2020-WW-000-04	40	0.0030	-	7470A	PPB	20 Nov 2020 11:46:39	U	SMPL
PB133116TB - 1	PB133097TB	-6	-0.0095	-	7470A	PPB	20 Nov 2020 11:48:55	U	SMPL
CCV66 - 1	CCV66	18444	5.0226	-	7470A	PPB	20 Nov 2020 11:51:11	U	SMPL
CCB66 - 1	CCB66	-62	-0.0248	-	7470A	PPB	20 Nov 2020 11:53:27	U	SMPL
L4786-02LX5 - 1		0	-0.0079	-	7470A	PPB	20 Nov 2020 11:55:42	U	SMPL
L4786-02A - 1		13806	3.7577	-	7470A	PPB	20 Nov 2020 11:57:57	U	SMPL
L4820-02LX5 - 1		-59	-0.0240	-	7470A	PPB	20 Nov 2020 12:00:13	U	SMPL
L4820-02A - 1		15239	4.1485	-	7470A	PPB	20 Nov 2020 12:02:28	U	SMPL
CCV67 - 1	CCV67	18330	4.9916	-	7470A	PPB	20 Nov 2020 12:04:44	U	SMPL
CCB67 - 1	CCB67	-83	-0.0305	-	7470A	PPB	20 Nov 2020 12:07:00	U	SMPL