



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112074

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/20/2020	13:10
2	CAL2	1	Water	NA	NA	20.1	7.00	11/20/2020	13:11
3	CAL3	1	Water	NA	NA	20.1	10.04	11/20/2020	13:12
4	ICV	1	Water	NA	NA	20.1	7.01	11/20/2020	13:15
5	CCV1	1	Water	NA	NA	20.1	2.01	11/20/2020	13:16
6	L4851-01	1	Water	NA	NA	18.8	5.03	11/20/2020	13:19
7	L4851-01DUP	1	Water	NA	NA	18.9	5.04	11/20/2020	13:20
8	CCV2	1	Water	NA	NA	20.1	12.02	11/20/2020	13:22

11207h

Date: 11-20-2020 12:58:14

Date/Time 11.20.2020 14:20

Received by: JD

Relinquished by: Cherry

Page 1 of 1

Date/Time 11/02/02 13.00
Received by: J. L. L. L. L.
Relinquished by: J. L. L. L.