

LB112075


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54348



A= 0.0000e+000

B= 1.8634e-004

C= 2.9547e-002

Rho= 0.9995853

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	40	0.000	40				
0.2	0.200	0.221	0.021	1030	0.0 %	1030				
2.5	2.500	2.549	0.049	13519	0.0 %	13519				
5.0	5.000	4.913	-0.087	26206	0.0 %	26206				
7.5	7.500	7.323	-0.177	39140	0.0 %	39140				
10.0	10.000	10.157	0.157	54348	0.0 %	54348				

LB112075 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Nov 2020 13:57:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	40	-	0.00007471B	PPB		20 Nov 2020 14:00:24	S	Std
0.2 - 1	SD 2	1030	-	0.20007471B	PPB		20 Nov 2020 14:02:41	S	Std
2.5 - 1	SD 1	13519	-	2.50007471B	PPB		20 Nov 2020 14:04:58	S	Std
5.0 - 1	SD	26206	-	5.00007471B	PPB		20 Nov 2020 14:07:14	S	Std
7.5 - 1	SD 1	39140	-	7.50007471B	PPB		20 Nov 2020 14:09:31	S	Std
10.0 - 1	SD	54348	-	10.00007471B	PPB		20 Nov 2020 14:17:04	S	Std
ICV14 - 1	ICV14	19957	3.7484	-7471B	PPB		20 Nov 2020 14:19:45	U	SMPL
ICB14 - 1	ICB14	-52	0.0199	-7471B	PPB		20 Nov 2020 14:22:00	U	SMPL
CCV68 - 1	CCV68	26196	4.9110	-7471B	PPB		20 Nov 2020 14:24:15	U	SMPL
CCB68 - 1	CCB68	-54	0.0195	-7471B	PPB		20 Nov 2020 14:26:30	U	SMPL
CRA - 1	CRA	987	0.2135	-7471B	PPB		20 Nov 2020 14:28:46	U	SMPL
HighStd - 1	HighStd	54421	10.1705	-7471B	PPB		20 Nov 2020 14:31:02	U	SMPL
ChkStd - 1	ChkStd	37451	7.0083	-7471B	PPB		20 Nov 2020 14:33:17	U	SMPL
PB133149BL - 1	PBS	-119	0.0074	-7471B	PPB		20 Nov 2020 14:35:34	U	SMPL
PB133149BS - 1	LCSS	18561	3.4883	-7471B	PPB		20 Nov 2020 14:37:51	U	SMPL
L4808-01 - 1	11931	1829	0.3704	-7471B	PPB		20 Nov 2020 14:40:07	U	SMPL
L4809-01 - 1	MS-CON-A	1932	0.3896	-7471B	PPB		20 Nov 2020 14:42:24	U	SMPL
L4809-03 - 1	MS-CON-B	4140	0.8010	-7471B	PPB		20 Nov 2020 14:44:40	U	SMPL
L4809-05 - 1	MS-CON-C	333	0.0916	-7471B	PPB		20 Nov 2020 14:46:57	U	SMPL
L4809-07 - 1	MS-CON-D	146	0.0568	-7471B	PPB		20 Nov 2020 14:49:12	U	SMPL
CCV69 - 1	CCV69	25751	4.8281	-7471B	PPB		20 Nov 2020 14:51:28	U	SMPL
CCB69 - 1	CCB69	-82	0.0143	-7471B	PPB		20 Nov 2020 14:53:43	U	SMPL
L4809-09 - 1	MS-CON-E	3373	0.6581	-7471B	PPB		20 Nov 2020 14:55:59	U	SMPL
L4809-11 - 1	MS-CON-F	414	0.1067	-7471B	PPB		20 Nov 2020 14:58:14	U	SMPL
L4810-01 - 1	TR-04-111920	2313	0.4606	-7471B	PPB		20 Nov 2020 15:00:29	U	SMPL
L4832-01 - 1	R-LA-33-2-WC	2027	0.4073	-7471B	PPB		20 Nov 2020 15:02:45	U	SMPL
L4832-01DUP - 1	R-LA-33-2-WCDUP	1994	0.4011	-7471B	PPB		20 Nov 2020 15:05:00	U	SMPL
L4832-01MS - 1	R-LA-33-2-WCMS	21406	4.0184	-7471B	PPB		20 Nov 2020 15:07:16	U	SMPL
L4832-01MSD - 1	R-LA-33-2-WCMSD	21129	3.9668	-7471B	PPB		20 Nov 2020 15:09:32	U	SMPL
L4832-01LX5 - 1		342	0.0933	-7471B	PPB		20 Nov 2020 15:11:49	U	SMPL
L4832-01A - 1		22467	4.2161	-7471B	PPB		20 Nov 2020 15:14:06	U	SMPL
CCV70 - 1	CCV70	25893	4.8545	-7471B	PPB		20 Nov 2020 15:16:22	U	SMPL
CCB70 - 1	CCB70	-79	0.0148	-7471B	PPB		20 Nov 2020 15:18:37	U	SMPL