



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112076

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	11/20/2020	12:05
2	L4826-02		1	8	100.00	>212.0	>212.0	11/20/2020	12:35
3	L4826-02DUP		1	8	100.00	>212.0	>212.0	11/20/2020	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 112076

WorkList Name : FP112020 WorkList ID : 144343 Department : Wet-Chemistry Date : 11-20-2020 09:00:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2020	Water	L4826-02	Flash Point	Cool 4 deg C	ARCA05	H33	PB-2020-WW-000-04	11/18/2020	1010A

Date/Time 11/20/20 11:30
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 11/20/20 13:15
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)