



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112083

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	11/20/2020	15:20
2	CAL2	1	Water	NA	NA	21.2	7.01	11/20/2020	15:21
3	CAL3	1	Water	NA	NA	21.2	10.03	11/20/2020	15:22
4	ICV	1	Water	NA	NA	21.3	7.00	11/20/2020	15:25
5	CCV1	1	Water	NA	NA	21.2	2.01	11/20/2020	15:26
6	L4845-04	1	Solid	20.03	20	21.3	7.48	11/20/2020	15:29
7	L4858-03	1	Solid	20.02	20	21.8	7.35	11/20/2020	15:30
8	L4858-04	1	Solid	20.01	20	22.1	8.22	11/20/2020	15:32
9	L4858-04DUP	1	Solid	20.03	20	22.2	8.23	11/20/2020	15:34
10	CCV2	1	Water	NA	NA	21.3	12.02	11/20/2020	15:35

18112083

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity I4858 WorkList ID : 144373 Department : Wet-Chemistry Date : 11-20-2020 15:01:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2020	Solid	L4845-04	Corrosivity	Cool 4 deg C	PSEG03	H33	PH-TP	11/20/2020	9045D
11/19/2020	Solid	L4858-03	Corrosivity	Cool 4 deg C	NOBI03	H51	PHC-340EM-012	11/19/2020	9045D
11/19/2020	Solid	L4858-04	Corrosivity	Cool 4 deg C	NOBI03	H51	PHC-340EM-013	11/19/2020	9045D

Date/Time 11-20-2020 15:10
Received by: J. (wet chem)
Relinquished by: CP (SM)

Date/Time 11-20-2020 17:30
Received by: CP (SM)
Relinquished by: J. (wet chem)