


LB112091

7471B

INSTRUMENT ID: CVI



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	384	0.000	384				
0.2	0.200	0.231	0.031	1594	0.0 %	1594				
2.5	2.500	2.507	0.007	14100	0.0 %	14100				
5.0	5.000	4.888	-0.112	27183	0.0 %	27183				
7.5	7.500	7.544	0.044	41776	0.0 %	41776				
10.0	10.000	10.019	0.019	55376	0.0 %	55376				

LB112091 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 23 Nov 2020 10:12:21						
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type	
0.0 - 1	50	384	-	0.0000	7471B	PPB	23 Nov 2020 10:16:04	S	Std	
0.2 - 1	50.2	1594	-	0.2000	7471B	PPB	23 Nov 2020 10:18:21	S	Std	
2.5 - 1	52.5	14100	-	2.5000	7471B	PPB	23 Nov 2020 10:20:38	S	Std	
5.0 - 1	55	27183	-	5.0000	7471B	PPB	23 Nov 2020 10:22:55	S	Std	
7.5 - 1	57.5	41776	-	7.5000	7471B	PPB	23 Nov 2020 10:25:11	S	Std	
10.0 - 1	518	55376	-	10.0000	7471B	PPB	23 Nov 2020 10:27:28	S	Std	
ICV15 - 1	ICV15	21747	3.8988	-	7471B	PPB	23 Nov 2020 10:30:13	U	SMPL	
ICB15 - 1	ICB15	493	0.0307	-	7471B	PPB	23 Nov 2020 10:32:27	U	SMPL	
CCV73 - 1	CCV73	28413	5.1119	-	7471B	PPB	23 Nov 2020 10:34:42	U	SMPL	
CCB73 - 1	CCB73	476	0.0276	-	7471B	PPB	23 Nov 2020 10:36:57	U	SMPL	
CRA - 1	CRA	1577	0.2280	-	7471B	PPB	23 Nov 2020 10:39:13	U	SMPL	
HighStd - 1	HighStd	56277	10.1830	-	7471B	PPB	23 Nov 2020 10:41:28	U	SMPL	
ChkStd - 1	ChkStd	40500	7.3117	-	7471B	PPB	23 Nov 2020 10:43:43	U	SMPL	
PB133164BL - 1	PBS	354	0.0054	-	7471B	PPB	23 Nov 2020 10:46:00	U	SMPL	
PB133164BS - 1	LCSS	21240	3.8065	-	7471B	PPB	23 Nov 2020 11:00:08	U	SMPL	
L4835-01 - 1	OF-3-20201119	16734	2.9864	-	7471B	PPB	23 Nov 2020 11:02:23	U	SMPL	
L4835-01DUP - 1	OF-3-20201119DUP	16295	2.9065	-	7471B	PPB	23 Nov 2020 11:04:40	U	SMPL	
L4835-02 - 1	L4835-01MS	37381	6.7441	-	7471B	PPB	23 Nov 2020 11:06:57	U	SMPL	
L4835-03 - 1	L4835-01MSD	38977	7.0345	-	7471B	PPB	23 Nov 2020 11:09:13	U	SMPL	
L4835-04 - 1	DUP03-20201119	21233	3.8052	-	7471B	PPB	23 Nov 2020 11:11:30	U	SMPL	
CCV74 - 1	CCV74	27499	4.9456	-	7471B	PPB	23 Nov 2020 11:13:47	U	SMPL	
CCB74 - 1	CCB74	451	0.0230	-	7471B	PPB	23 Nov 2020 11:16:03	U	SMPL	
L4836-01 - 1	SB-01	59749	10.8149	-	7471B	PPB	23 Nov 2020 11:18:18	U	SMPL	
L4836-02 - 1	SB-02	25057	4.5012	-	7471B	PPB	23 Nov 2020 11:20:33	U	SMPL	
L4836-03 - 1	SB-03	35557	6.4121	-	7471B	PPB	23 Nov 2020 11:22:51	U	SMPL	
L4836-04 - 1	SB-04	34336	6.1899	-	7471B	PPB	23 Nov 2020 11:25:08	U	SMPL	
L4836-05 - 1	SB-05	70781	12.8226	-	7471B	PPB	23 Nov 2020 11:27:25	U	SMPL	
L4836-06 - 1	SB-06	20830	3.7319	-	7471B	PPB	23 Nov 2020 11:29:43	U	SMPL	
L4836-07 - 1	SB-07	23381	4.1961	-	7471B	PPB	23 Nov 2020 11:32:01	U	SMPL	
L4836-08 - 1	SB-08	4683	0.7932	-	7471B	PPB	23 Nov 2020 11:34:18	U	SMPL	
L4836-09 - 1	SL-02	5014	0.8535	-	7471B	PPB	23 Nov 2020 11:36:36	U	SMPL	
L4836-10 - 1	SL-01	5068	0.8633	-	7471B	PPB	23 Nov 2020 11:38:52	U	SMPL	
CCV75 - 1	CCV75	27525	4.9503	-	7471B	PPB	23 Nov 2020 11:41:07	U	SMPL	
CCB75 - 1	CCB75	375	0.0092	-	7471B	PPB	23 Nov 2020 11:43:23	U	SMPL	
PB133165BL - 1	PBS	509	0.0336	-	7471B	PPB	23 Nov 2020 11:45:39	U	SMPL	
PB133165BS - 1	LCSS	21168	3.7934	-	7471B	PPB	23 Nov 2020 11:50:03	U	SMPL	
L4592-01 - 1	EO-01-112020	1648	0.2409	-	7471B	PPB	23 Nov 2020 11:52:18	U	SMPL	
L4592-01DUP - 1	EO-01-112020DUP	1735	0.2567	-	7471B	PPB	23 Nov 2020 11:54:34	U	SMPL	
L4592-01MS - 1	EO-01-112020MS	19485	3.4871	-	7471B	PPB	23 Nov 2020 11:56:49	U	SMPL	
L4592-01MSD - 1	EO-01-112020MSD	20883	3.7415	-	7471B	PPB	23 Nov 2020 12:01:42	U	SMPL	
L4838-02 - 1	KENS-B2-111920-0-2	3847	0.6411	-	7471B	PPB	23 Nov 2020 12:04:00	U	SMPL	
L4838-04 - 1	KENS-B2-111920-6-8	1035	0.1293	-	7471B	PPB	23 Nov 2020 12:06:18	U	SMPL	
L4842-01 - 1	11620	593	0.0489	-	7471B	PPB	23 Nov 2020 12:08:34	U	SMPL	
L4844-01 - 1	OR-02-112020	1539	0.2210	-	7471B	PPB	23 Nov 2020 12:10:50	U	SMPL	
CCV76 - 1	CCV76	26358	4.7379	-	7471B	PPB	23 Nov 2020 12:13:06	U	SMPL	
CCB76 - 1	CCB76	273	-0.0094	-	7471B	PPB	23 Nov 2020 12:15:21	U	SMPL	
L4846-01 - 1	RCA-EDISON-3	1233	0.1654	-	7471B	PPB	23 Nov 2020 12:17:38	U	SMPL	
L4848-01 - 1	TR5-(1-4)-COMP	8006	1.3980	-	7471B	PPB	23 Nov 2020 12:19:53	U	SMPL	
L4857-01 - 1	304EM-STOCKPILE	10955	1.9347	-	7471B	PPB	23 Nov 2020 12:22:09	U	SMPL	
L4857-02 - 1	304EM-STOCKPILE-EAST	3697	0.6138	-	7471B	PPB	23 Nov 2020 12:24:26	U	SMPL	
L4857-03 - 1	304EM-STOCKPILE-WEST	10026	1.7656	-	7471B	PPB	23 Nov 2020 12:26:43	U	SMPL	
L4858-01 - 1	PHC-340EM-012	40736	7.3546	-	7471B	PPB	23 Nov 2020 12:28:59	U	SMPL	
L4858-02 - 1	PHC-340EM-013	33890	6.1087	-	7471B	PPB	23 Nov 2020 12:31:16	U	SMPL	
L4868-02 - 1	KENS-B3-112020-0-2	1513	0.2163	-	7471B	PPB	23 Nov 2020 12:33:35	U	SMPL	
L4868-04 - 1	KENS-B3-112020-8-10	1104	0.1419	-	7471B	PPB	23 Nov 2020 12:35:54	U	SMPL	
L4868-06 - 1	KENS-B3-112020-BD	2361	0.3706	-	7471B	PPB	23 Nov 2020 12:38:10	U	SMPL	
CCV77 - 1	CCV77	26686	4.7976	-	7471B	PPB	23 Nov 2020 12:40:25	U	SMPL	
CCB77 - 1	CCB77	294	-0.0055	-	7471B	PPB	23 Nov 2020 12:42:42	U	SMPL	
L4835-01LX5 - 1		3429	0.5650	-	7471B	PPB	23 Nov 2020 12:44:57	U	SMPL	
L4835-01A - 1		35899	6.4743	-	7471B	PPB	23 Nov 2020 12:47:13	U	SMPL	
L4592-01LX5 - 1		504	0.0327	-	7471B	PPB	23 Nov 2020 12:49:28	U	SMPL	
L4592-01A - 1		21514	3.8564	-	7471B	PPB	23 Nov 2020 12:51:45	U	SMPL	
L4836-01DLX2 - 1		30292	5.4539	-	7471B	PPB	23 Nov 2020 12:54:01	U	SMPL	
L4836-05DLX2 - 1		35129	6.3342	-	7471B	PPB	23 Nov 2020 12:56:18	U	SMPL	
CCV78 - 1	CCV78	27294	4.9083	-	7471B	PPB	23 Nov 2020 12:58:35	U	SMPL	
CCB78 - 1	CCB78	459	0.0245	-	7471B	PPB	23 Nov 2020 13:00:54	U	SMPL	