



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB112099

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	11/23/2020	14:15
2	CAL2	1	Water	NA	NA	21.2	7.00	11/23/2020	14:16
3	CAL3	1	Water	NA	NA	21.2	10.03	11/23/2020	14:17
4	ICV	1	Water	NA	NA	21.2	7.00	11/23/2020	14:18
5	CCV1	1	Water	NA	NA	21.3	2.02	11/23/2020	14:19
6	L4840-03	1	Water	NA	NA	23.4	6.93	11/23/2020	14:22
7	L4840-04	1	Water	NA	NA	23.2	6.66	11/23/2020	14:24
8	L4840-05	1	Water	NA	NA	23.1	9.18	11/23/2020	14:25
9	L4840-06	1	Water	NA	NA	23.3	5.57	11/23/2020	14:26
10	L4841-01	1	Water	NA	NA	23.3	6.41	11/23/2020	14:30
11	L4841-02	1	Water	NA	NA	23.4	7.91	11/23/2020	14:32
12	L4841-03	1	Water	NA	NA	23.1	9.50	11/23/2020	14:33
13	L4842-03	1	Water	NA	NA	23.4	4.50	11/23/2020	14:35
14	L4842-04	1	Water	NA	NA	23.2	8.15	11/23/2020	14:36
15	L4842-06	1	Water	NA	NA	23.3	7.38	11/23/2020	14:38
16	CCV2	1	Water	NA	NA	21.1	12.02	11/23/2020	14:39
17	L4842-06DUP	1	Water	NA	NA	23.4	7.39	11/23/2020	14:40
18	CCV3	1	Water	NA	NA	21.2	2.02	11/23/2020	14:41

WORKLIST(Hardcopy Internal Chain)

283112099

WorkList Name : PH L4871 W

WorkList ID : 144424

Department : Wet-Chemistry

Date : 11-23-2020 14:04:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2020	Water	L4840-03	pH	Cool 4 deg C	PSEG03	H41	16857	11/20/2020	9040C
11/20/2020	Water	L4840-04	pH	Cool 4 deg C	PSEG03	H41	41612	11/20/2020	9040C
11/20/2020	Water	L4840-05	pH	Cool 4 deg C	PSEG03	H41	16858	11/20/2020	9040C
11/20/2020	Water	L4840-06	pH	Cool 4 deg C	PSEG03	H41	30738	11/20/2020	9040C
11/20/2020	Water	L4841-01	pH	Cool 4 deg C	PSEG03	H33	40741	11/20/2020	9040C
11/20/2020	Water	L4841-02	pH	Cool 4 deg C	PSEG03	H33	16855	11/20/2020	9040C
11/20/2020	Water	L4841-03	pH	Cool 4 deg C	PSEG03	H33	16881	11/20/2020	9040C
11/20/2020	Water	L4842-03	pH	Cool 4 deg C	PSEG03	H41	16846-41629	11/20/2020	9040C
11/20/2020	Water	L4842-04	pH	Cool 4 deg C	PSEG03	H41	40642	11/20/2020	9040C
11/20/2020	Water	L4842-06	pH	Cool 4 deg C	PSEG03	H41	11920	11/20/2020	9040C

Date/Time

Received by:

Relinquished by:

11/23/2020 14:10

JO (Wet Chem)

KS (Wet Chem)

Date/Time

Received by:

Relinquished by:

11/23/2020 16:19

KS (Wet Chem)

JO (Wet Chem)