



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112100

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	11/23/2020	14:15
2	CAL2	1	Water	NA	NA	21.2	7.00	11/23/2020	14:16
3	CAL3	1	Water	NA	NA	21.2	10.03	11/23/2020	14:17
4	ICV	1	Water	NA	NA	21.2	7.00	11/23/2020	14:18
5	CCV1	1	Water	NA	NA	21.3	2.02	11/23/2020	14:19
6	L4840-01	1	Solid	20.03	20	21.5	5.29	11/23/2020	14:50
7	L4840-02	1	Solid	20.05	20	23.0	5.47	11/23/2020	14:51
8	L4871-01	1	Solid	20.02	20	23.1	8.29	11/23/2020	14:53
9	CCV2	1	Water	NA	NA	21.2	12.03	11/23/2020	15:00

VB 112100

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4872 S WorkList ID : 144425 Department : Wet-Chemistry Date : 11-23-2020 14:05:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/20/2020	Solid	L4840-01	pH	Cool 4 deg C	PSEG03	H41	30739	11/20/2020	9045D
11/20/2020	Solid	L4840-02	pH	Cool 4 deg C	PSEG03	H41	30740	11/20/2020	9045D
11/23/2020	Solid	L4871-01	pH	Cool 4 deg C	PSEG03	H51	32783	11/23/2020	9045D

Date/Time 11-23-2020 14:10
Received by: Jo (Wetchem)
Relinquished by: KS (Wetchem)

Date/Time 11-23-2020 26:10
Received by: KS (Wetchem)
Relinquished by: Jo (Wetchem) Chen