

LB112102


SFAM01.0

INSTRUMENT ID : CV1

Linear

 μ Abs.:

38892



A= 0.0000e+000

B= 2.5717e-004

C= 3.6976e-003

Rho= 0.9999677

Accept=Accepted

slope
Y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.024	0.024	78	0.000	78					11
0.05	0.050					244					8
0.20	0.200	0.215	0.015	822	0.0 %	822					-3
2.50	2.500	2.436	-0.064	9459	0.0 %	9459					0
5.00	5.000	5.020	0.020	19505	0.0 %	19505					0
7.50	7.500	7.500	-0.000	29149	0.0 %	29149					0
10.0	10.000	10.005	0.005	38892	0.0 %	38892					0

LB112102
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	78	-		0 SFAM01.0	PPB	11/23/2020 14:44	Std
	0.2 S0.2	822	-		0.2 SFAM01.0	PPB	11/23/2020 14:47	Std
	2.5 S2.5	9459	-		2.5 SFAM01.0	PPB	11/23/2020 14:49	Std
	5 S5.0	19505	-		5 SFAM01.0	PPB	11/23/2020 14:51	Std
	7.5 S7.5	29149	-		7.5 SFAM01.0	PPB	11/23/2020 14:54	Std
	10 S10.0	38892	-		10 SFAM01.0	PPB	11/23/2020 15:01	Std
ICV	ICV	14075	3.6233 -		SFAM01.0	PPB	11/23/2020 15:04	SMPL
ICB	ICB	-41	-0.0068 -		SFAM01.0	PPB	11/23/2020 15:06	SMPL
CCV079	CCV079	19456	5.0071 -		SFAM01.0	PPB	11/23/2020 15:09	SMPL
CCB079	CCB079	-33	-0.0048 -		SFAM01.0	PPB	11/23/2020 15:11	SMPL
PB133193BL	PBW319	53	0.0173 -		SFAM01.0	PPB	11/23/2020 15:13	SMPL
L4823-07	X2Q51	13411	3.4525 -		SFAM01.0	PPB	11/23/2020 15:15	SMPL
L4823-08	X2Q51D	13531	3.4834 -		SFAM01.0	PPB	11/23/2020 15:20	SMPL
L4823-09	X2Q51S	17282	4.448 -		SFAM01.0	PPB	11/23/2020 15:22	SMPL
CCV080	CCV080	19526	5.0251 -		SFAM01.0	PPB	11/23/2020 15:24	SMPL
CCB080	CCB080	-37	-0.0058 -		SFAM01.0	PPB	11/23/2020 15:27	SMPL