



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112109

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/23/2020	11:40
2	L4840-03		1	8	100.00	>212.0	>212.0	11/23/2020	12:10
3	L4840-03DUP		1	8	100.00	>212.0	>212.0	11/23/2020	12:35
4	L4840-04		1	9	100.00	>212.0	>212.0	11/23/2020	13:05
5	L4840-05		1	8	100.00	>212.0	>212.0	11/23/2020	13:32
6	L4840-06		1	9	100.00	>212.0	>212.0	11/23/2020	13:58
7	L4841-01		1	7	100.00	>212.0	>212.0	11/23/2020	14:28
8	L4841-02		1	8	100.00	>212.0	>212.0	11/23/2020	14:55
9	L4841-03		1	8	100.00	>212.0	>212.0	11/23/2020	15:20
10	L4842-03		1	8	100.00	>212.0	>212.0	11/23/2020	15:45
11	L4842-04		1	9	100.00	>212.0	>212.0	11/23/2020	16:10
12	L4842-06		1	8	100.00	>212.0	>212.0	11/23/2020	16:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB112109

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp112320 WorkList ID : 144409 Department : Wet-Chemistry Date : 11-23-2020 09:02:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/27/2020	Water	L4840-03	Flash Point	Cool 4 deg C	PSEG03	H41	16857	11/20/2020	1010A
11/27/2020	Water	L4840-04	Flash Point	Cool 4 deg C	PSEG03	H41	41612	11/20/2020	1010A
11/27/2020	Water	L4840-05	Flash Point	Cool 4 deg C	PSEG03	H41	16858	11/20/2020	1010A
11/27/2020	Water	L4840-06	Flash Point	Cool 4 deg C	PSEG03	H41	30738	11/20/2020	1010A
11/27/2020	Water	L4841-01	Flash Point	Cool 4 deg C	PSEG03	H33	40741	11/20/2020	1010A
11/27/2020	Water	L4841-02	Flash Point	Cool 4 deg C	PSEG03	H33	16855	11/20/2020	1010A
11/27/2020	Water	L4841-03	Flash Point	Cool 4 deg C	PSEG03	H33	16881	11/20/2020	1010A
11/27/2020	Water	L4842-03	Flash Point	Cool 4 deg C	PSEG03	H41	16846-41629	11/20/2020	1010A
11/27/2020	Water	L4842-04	Flash Point	Cool 4 deg C	PSEG03	H41	40642	11/20/2020	1010A
11/27/2020	Water	L4842-06	Flash Point	Cool 4 deg C	PSEG03	H41	11920	11/20/2020	1010A

Date/Time 11/23/2020 11:15
Received by: KS (Wetchem)
Relinquished by: KS (Wetchem)

Date/Time 11/23/2020 16:40
Received by: KS (Wetchem)
Relinquished by: KS (Wetchem)