



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112114

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	11/23/2020	16:15
2	CAL2	1	Water	NA	NA	21.2	7.00	11/23/2020	16:16
3	CAL3	1	Water	NA	NA	21.2	10.03	11/23/2020	16:17
4	ICV	1	Water	NA	NA	21.2	6.99	11/23/2020	16:19
5	CCV1	1	Water	NA	NA	21.2	2.02	11/23/2020	16:20
6	L4597-04	1	Solid	20.03	20	21.9	7.21	11/23/2020	16:23
7	L4597-04DUP	1	Solid	20.04	20	21.8	7.22	11/23/2020	16:24
8	L4882-04	1	Solid	20.01	20	20.6	11.30	11/23/2020	16:25
9	L4882-06	1	Solid	20.03	20	21.7	10.09	11/23/2020	16:26
10	CCV2	1	Water	NA	NA	21.2	12.02	11/23/2020	16:28

VB 112114

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity I4879 WorkList ID : 144428 Department : Wet-Chemistry Date : 11-23-2020 16:58:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/23/2020	Solid	L4597-04	Corrosivity	Cool 4 deg C	PSEG05	H53	BU-03-112320	11/23/2020	9045D
11/23/2020	Solid	L4882-04	Corrosivity	Cool 4 deg C	ENTE01	H51	PAV-B48-112320-0-10	11/23/2020	9045D
11/23/2020	Solid	L4882-06	Corrosivity	Cool 4 deg C	ENTE01	H51	PAV-B48-112320-10-16.5	11/23/2020	9045D

Date/Time 11-23-2020 16:10
 Received by: J. West
 Relinquished by: J. West

Date/Time 11-23-2020 17:35
 Received by: J. West
 Relinquished by: J. West