


SFAM01.0

LB112122
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44151



A= 0.0000e+000

B= 2.2457e-004

C= -1.8236e-003

Rho= 0.9998250

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	0.009	0.009	46	0.000	46					1
0.05	0.050					244					1
0.20	0.200	0.204	0.004	915	0.0 %	915					2
2.50	2.500	2.449	-0.051	10912	0.0 %	10912					-2
5.00	5.000	4.991	-0.009	22233	0.0 %	22233					0
7.50	7.500	7.635	0.135	34007	0.0 %	34007					2
10.0	10.000	9.913	-0.087	44151	0.0 %	44151					-1

LB112122

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	46	-		0 ISM02.4	PPB	11/24/2020 14:20	Std
	0.2 S0.2	915	-		0.2 ISM02.4	PPB	11/24/2020 14:22	Std
	2.5 S2.5	10912	-		2.5 ISM02.4	PPB	11/24/2020 14:25	Std
	5 S5.0	22233	-		5 ISM02.4	PPB	11/24/2020 14:27	Std
	7.5 S7.5	34007	-		7.5 ISM02.4	PPB	11/24/2020 14:29	Std
	10 S10.0	44151	-		10 ISM02.4	PPB	11/24/2020 14:32	Std
ICV	ICV	17710	3.9753 -		ISM02.4	PPB	11/24/2020 14:34	SMPL
ICB	ICB	-51	-0.0133 -		ISM02.4	PPB	11/24/2020 14:37	SMPL
CCV090	CCV090	22084	4.9575 -		ISM02.4	PPB	11/24/2020 14:39	SMPL
CCB090	CCB090	-66	-0.0166 -		ISM02.4	PPB	11/24/2020 14:41	SMPL
PB133237BL	PBS237	6	-0.0005 -		ISM02.4	PPB	11/24/2020 14:43	SMPL
L4823-19	X2Q55	18032	4.0476 -		ISM02.4	PPB	11/24/2020 14:48	SMPL
L4823-20	X2Q55D	18012	4.0431 -		ISM02.4	PPB	11/24/2020 14:50	SMPL
L4823-21	X2Q55S	28368	6.3687 -		ISM02.4	PPB	11/24/2020 15:02	SMPL
CCV091	CCV091	21727	4.8774 -		ISM02.4	PPB	11/24/2020 15:04	SMPL
CCB091	CCB091	-41	-0.011 -		ISM02.4	PPB	11/24/2020 15:06	SMPL