



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112129

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	11/24/2020	17:25
2	CAL2	1	Water	NA	NA	21.2	7.00	11/24/2020	17:26
3	CAL3	1	Water	NA	NA	21.2	10.03	11/24/2020	17:27
4	ICV	1	Water	NA	NA	21.2	7.00	11/24/2020	17:29
5	CCV1	1	Water	NA	NA	20.2	2.01	11/24/2020	17:30
6	L4884-01	1	Solid	20.02	20	20.1	7.26	11/24/2020	17:32
7	L4891-04	1	Solid	20.03	20	21.5	9.40	11/24/2020	17:35
8	L4891-06	1	Solid	20.01	20	21.6	9.33	11/24/2020	17:37
9	L4891-06DUP	1	Solid	20.02	20	21.7	9.34	11/24/2020	17:38
10	CCV2	1	Water	NA	NA	21.3	12.02	11/24/2020	17:39

VB 112129

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity I4591 WorkList ID : 144463 Department : Wet-Chemistry Date : 11-24-2020 17:20:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/24/2020	Solid	L4884-01	Corrosivity	Cool 4 deg C	PSEG03	J11	FS-OD	11/24/2020	9045D
11/24/2020	Solid	L4891-04	Corrosivity	Cool 4 deg C	ENTE01	J12	PAV-B3-112420-0-10	11/24/2020	9045D
11/24/2020	Solid	L4891-06	Corrosivity	Cool 4 deg C	ENTE01	J12	PAV-B3-112420-10-11.5	11/24/2020	9045D

Date/Time 11.24.2020 17:22
Received by: Jn / 609
Relinquished by: Jn / 609

Date/Time 11.24.2020 18:00
Received by: Jn / 609
Relinquished by: Jn / 609