


LB112134

7470A

INSTRUMENT

 μ Abs.:

36569



Linear

A= 0.0000e+000

B= 2.7378e-004

C= 6.0929e-003

Rho= 0.9998618

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	-22	0.000	-22				
0.2	0.200	0.198	-0.002	702	0.0 %	702				
2.5	2.500	2.459	-0.041	8961	0.0 %	8961				
5.0	5.000	5.113	0.113	18653	0.0 %	18653				
7.5	7.500	7.411	-0.089	27048	0.0 %	27048				
10.0	10.000	10.018	0.018	36569	0.0 %	36569				

LB112134

Method: 7470A	Operator: Admin	Date of Analysis: 25 Nov 2020 09:03:05							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-22	-	0.0000	7470A	PPB	25 Nov 2020 09:12:11	S	Std
0.2 - 1	50.2	702	-	0.2000	7470A	PPB	25 Nov 2020 09:14:29	S	Std
2.5 - 1	52.5	8961	-	2.5000	7470A	PPB	25 Nov 2020 09:16:46	S	Std
5.0 - 1	55	18653	-	5.0000	7470A	PPB	25 Nov 2020 09:19:02	S	Std
7.5 - 1	57.5	27048	-	7.5000	7470A	PPB	25 Nov 2020 09:21:19	S	Std
10.0 - 1	510	36569	-	10.0000	7470A	PPB	25 Nov 2020 09:23:35	S	Std
ICV17 - 1	ICV17	13697	3.7561	-	7470A	PPB	25 Nov 2020 09:26:19	U	SMPL
ICB17 - 1	ICB17	-78	-0.0153	-	7470A	PPB	25 Nov 2020 09:28:34	U	SMPL
CCV92 - 1	CCV92	18591	5.0959	-	7470A	PPB	25 Nov 2020 09:30:49	U	SMPL
CCB92 - 1	CCB92	-76	-0.0147	-	7470A	PPB	25 Nov 2020 09:33:04	U	SMPL
CRA - 1	CRA	697	0.1969	-	7470A	PPB	25 Nov 2020 09:35:19	U	SMPL
HighStd - 1	HighStd	36120	9.8950	-	7470A	PPB	25 Nov 2020 09:37:34	U	SMPL
ChkStd - 1	ChkStd	25731	7.0507	-	7470A	PPB	25 Nov 2020 09:39:50	U	SMPL
PB133236BL - 1	PBW	-148	-0.0344	-	7470A	PPB	25 Nov 2020 09:42:07	U	SMPL
PB133236BS - 1	LCSW	13477	3.6958	-	7470A	PPB	25 Nov 2020 09:44:23	U	SMPL
L4597-04 - 1	BU-03-112320	-57	-0.0095	-	7470A	PPB	25 Nov 2020 09:46:40	U	SMPL
L4874-03 - 1	72-11933-1	-16	0.0017	-	7470A	PPB	25 Nov 2020 09:48:57	U	SMPL
L4879-01 - 1	P0102-CAS001-0002-01	2	0.0066	-	7470A	PPB	25 Nov 2020 09:51:13	U	SMPL
L4879-02 - 1	P0102-CAS001-0204-01	-15	0.0020	-	7470A	PPB	25 Nov 2020 09:53:30	U	SMPL
L4879-03 - 1	P0102-CAS001-0406-01	4	0.0072	-	7470A	PPB	25 Nov 2020 09:55:45	U	SMPL
CCV93 - 1	CCV93	18095	4.9602	-	7470A	PPB	25 Nov 2020 09:58:01	U	SMPL
CCB93 - 1	CCB93	-69	-0.0128	-	7470A	PPB	25 Nov 2020 10:00:16	U	SMPL
L4879-04 - 1	P0102-CAS002-0002-01	49	0.0195	-	7470A	PPB	25 Nov 2020 10:02:32	U	SMPL
L4879-05 - 1	P0102-CAS002-0204-01	42	0.0176	-	7470A	PPB	25 Nov 2020 10:04:47	U	SMPL
L4879-06 - 1	P0102-CAS002-0406-01	35	0.0157	-	7470A	PPB	25 Nov 2020 10:07:02	U	SMPL
L4879-07 - 1	P0122-CAS001-0002-01	39	0.0168	-	7470A	PPB	25 Nov 2020 10:09:18	U	SMPL
L4879-07DUP - 1	P0122-CAS001-0002-01DUP	27	0.0135	-	7470A	PPB	25 Nov 2020 10:11:34	U	SMPL
L4879-08 - 1	L4879-07MS	14087	3.8628	-	7470A	PPB	25 Nov 2020 10:13:50	U	SMPL
L4879-09 - 1	L4879-07MSD	13529	3.7101	-	7470A	PPB	25 Nov 2020 10:16:06	U	SMPL
L4879-10 - 1	P0122-CAS001-0002-02	-43	-0.0057	-	7470A	PPB	25 Nov 2020 10:18:24	U	SMPL
L4879-11 - 1	P0122-CAS001-0204-01	19	0.0113	-	7470A	PPB	25 Nov 2020 10:20:41	U	SMPL
L4879-12 - 1	P0122-CAS001-0406-01	14	0.0099	-	7470A	PPB	25 Nov 2020 10:22:56	U	SMPL
CCV94 - 1	CCV94	18473	5.0636	-	7470A	PPB	25 Nov 2020 10:25:12	U	SMPL
CCB94 - 1	CCB94	-68	-0.0125	-	7470A	PPB	25 Nov 2020 10:27:27	U	SMPL
L4882-04 - 1	PAV-B48-112320-0-10	16	0.0105	-	7470A	PPB	25 Nov 2020 10:29:43	U	SMPL
L4882-06 - 1	PAV-B48-112320-10-16.5	26	0.0132	-	7470A	PPB	25 Nov 2020 10:31:58	U	SMPL
PB133236TB - 1	PB133212TB	14	0.0099	-	7470A	PPB	25 Nov 2020 10:34:14	U	SMPL
L4879-07LX5 - 1		8	0.0083	-	7470A	PPB	25 Nov 2020 10:36:30	U	SMPL
L4879-07A - 1		14241	3.9050	-	7470A	PPB	25 Nov 2020 10:38:46	U	SMPL
CCV95 - 1	CCV95	18318	5.0212	-	7470A	PPB	25 Nov 2020 10:41:02	U	SMPL
CCB95 - 1	CCB95	-61	-0.0106	-	7470A	PPB	25 Nov 2020 10:43:18	U	SMPL