


7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	2	0.000	2				
0.2	0.200	0.208	0.008	1091	0.0 %	1091				
2.5	2.500	2.508	0.008	13922	0.0 %	13922				
5.0	5.000	4.901	-0.099	27276	0.0 %	27276				
7.5	7.500	7.599	0.099	42335	0.0 %	42335				
10.0	10.000	9.971	-0.029	55570	0.0 %	55570				

LB112140 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 25 Nov 2020 12:20:03					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	2	-	0.0000	7471B	PPB	25 Nov 2020 12:23:15	S	Std
0.2 - 1	50.2	1091	-	0.2000	7471B	PPB	25 Nov 2020 12:25:32	S	Std
2.5 - 1	50.5	13922	-	2.5000	7471B	PPB	25 Nov 2020 12:27:49	S	Std
5.0 - 1	55	27276	-	5.0000	7471B	PPB	25 Nov 2020 12:30:06	S	Std
7.5 - 1	57.5	42335	-	7.5000	7471B	PPB	25 Nov 2020 12:32:23	S	Std
10.0 - 1	510	55570	-	10.0000	7471B	PPB	25 Nov 2020 12:37:14	S	Std
ICV18 - 1	ICV18	20254	3.6423	-	7471B	PPB	25 Nov 2020 12:39:53	U	SMPL
ICB18 - 1	ICB18	-118	-0.0084	-	7471B	PPB	25 Nov 2020 12:42:08	U	SMPL
CCV96 - 1	CCV96	27379	4.9191	-	7471B	PPB	25 Nov 2020 12:44:23	U	SMPL
CCB96 - 1	CCB96	-100	-0.0052	-	7471B	PPB	25 Nov 2020 12:46:38	U	SMPL
CRA - 1	CRA	1070	0.2045	-	7471B	PPB	25 Nov 2020 12:48:54	U	SMPL
HighStd - 1	HighStd	56222	10.0879	-	7471B	PPB	25 Nov 2020 12:51:10	U	SMPL
ChkStd - 1	ChkStd	39406	7.0744	-	7471B	PPB	25 Nov 2020 12:53:25	U	SMPL
PB133235BL - 1	PBS	-128	-0.0102	-	7471B	PPB	25 Nov 2020 12:55:42	U	SMPL
PB133235BS - 1	LCSS	18996	3.4169	-	7471B	PPB	25 Nov 2020 12:57:59	U	SMPL
L4588-01 - 1	HD-02-112320	5613	1.0186	-	7471B	PPB	25 Nov 2020 13:00:16	U	SMPL
L4597-01 - 1	BU-03-112320	1321	0.2495	-	7471B	PPB	25 Nov 2020 13:02:33	U	SMPL
L4631-03 - 1	PAV-B11-110420-0-10	8480	1.5324	-	7471B	PPB	25 Nov 2020 13:04:50	U	SMPL
L4631-05 - 1	PAV-B11-110420-10-15.25	6378	1.1557	-	7471B	PPB	25 Nov 2020 13:07:07	U	SMPL
L4872-01 - 1	SU-04-112320	1155	0.2197	-	7471B	PPB	25 Nov 2020 13:09:23	U	SMPL
CCV97 - 1	CCV97	27042	4.8587	-	7471B	PPB	25 Nov 2020 13:11:38	U	SMPL
CCB97 - 1	CCB97	-62	0.0016	-	7471B	PPB	25 Nov 2020 13:13:54	U	SMPL
L4873-01 - 1	HR-01-112320	2668	0.4909	-	7471B	PPB	25 Nov 2020 13:16:09	U	SMPL
L4874-01 - 1	72-11933	6824	1.2356	-	7471B	PPB	25 Nov 2020 13:18:25	U	SMPL
L4875-01 - 1	TR-08-112320	1436	0.2701	-	7471B	PPB	25 Nov 2020 13:20:40	U	SMPL
L4880-02 - 1	KENS-B4-112320-0-2	2301	0.4251	-	7471B	PPB	25 Nov 2020 13:22:56	U	SMPL
L4880-02DUP - 1	KENS-B4-112320-0-2DUP	2123	0.3932	-	7471B	PPB	25 Nov 2020 13:25:12	U	SMPL
L4880-04 - 1	L4880-02MS	24582	4.4179	-	7471B	PPB	25 Nov 2020 13:27:28	U	SMPL
L4880-06 - 1	L4880-02MSD	21320	3.8334	-	7471B	PPB	25 Nov 2020 13:29:45	U	SMPL
L4880-08 - 1	KENS-B4-112320-6-8	651	0.1294	-	7471B	PPB	25 Nov 2020 13:32:02	U	SMPL
L4882-03 - 1	PAV-B48-112320-0-10	8875	1.6032	-	7471B	PPB	25 Nov 2020 13:34:19	U	SMPL
L4882-05 - 1	PAV-B48-112320-10-16.5	4147	0.7559	-	7471B	PPB	25 Nov 2020 13:36:34	U	SMPL
CCV98 - 1	CCV98	27431	4.9285	-	7471B	PPB	25 Nov 2020 13:38:50	U	SMPL
CCB98 - 1	CCB98	-71	0.0000	-	7471B	PPB	25 Nov 2020 13:41:05	U	SMPL
L4884-01 - 1	FS-OD	601	0.1205	-	7471B	PPB	25 Nov 2020 13:43:21	U	SMPL
L4885-01 - 1	OK-03-112420	1078	0.2059	-	7471B	PPB	25 Nov 2020 13:45:37	U	SMPL
L4891-03 - 1	PAV-B3-112420-0-10	21832	3.9251	-	7471B	PPB	25 Nov 2020 13:47:52	U	SMPL
L4891-05 - 1	PAV-B3-112420-10-11.5	9227	1.6663	-	7471B	PPB	25 Nov 2020 13:50:08	U	SMPL
L4893-01 - 1	R-LA-36-2-WC	2655	0.4885	-	7471B	PPB	25 Nov 2020 13:52:23	U	SMPL
L4894-01 - 1	R-LA-36-4-WC	1006	0.1930	-	7471B	PPB	25 Nov 2020 13:54:40	U	SMPL
L4880-02LX5 - 1		389	0.0825	-	7471B	PPB	25 Nov 2020 13:56:56	U	SMPL
L4880-02A - 1		23909	4.2973	-	7471B	PPB	25 Nov 2020 13:59:13	U	SMPL
CCV99 - 1	CCV99	27295	4.9041	-	7471B	PPB	25 Nov 2020 14:01:30	U	SMPL
CCB99 - 1	CCB99	-72	-0.0001	-	7471B	PPB	25 Nov 2020 14:03:45	U	SMPL