



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112143

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	11/25/2020	14:05
2	CAL2	1	Water	NA	NA	21.2	7.00	11/25/2020	14:06
3	CAL3	1	Water	NA	NA	21.2	10.03	11/25/2020	14:07
4	ICV	1	Water	NA	NA	21.2	7.00	11/25/2020	14:10
5	CCV1	1	Water	NA	NA	21.3	2.02	11/25/2020	14:11
6	L4895-01	1	Water	NA	NA	18.8	6.28	11/25/2020	14:13
7	L4895-01DUP	1	Water	NA	NA	18.9	6.29	11/25/2020	14:14
8	L4895-03	1	Water	NA	NA	17.9	6.95	11/25/2020	14:19
9	L4896-01	1	Water	NA	NA	14.2	7.72	11/25/2020	14:21
10	L4896-02	1	Water	NA	NA	14.0	7.74	11/25/2020	14:23
11	L4896-03	1	Water	NA	NA	14.7	7.75	11/25/2020	14:25
12	L4896-04	1	Water	NA	NA	15.0	7.73	11/25/2020	14:26
13	L4896-05	1	Water	NA	NA	14.8	7.74	11/25/2020	14:27
14	L4896-06	1	Water	NA	NA	14.7	7.72	11/25/2020	14:29
15	L4896-07	1	Water	NA	NA	15.2	7.77	11/25/2020	14:30
16	CCV2	1	Water	NA	NA	21.3	12.02	11/25/2020	14:32
17	L4896-08	1	Water	NA	NA	15.5	7.78	11/25/2020	14:35
18	L4896-09	1	Water	NA	NA	15.7	7.73	11/25/2020	14:36
19	L4896-10	1	Water	NA	NA	15.9	7.72	11/25/2020	14:38
20	CCV3	1	Water	NA	NA	21.3	2.01	11/25/2020	14:39

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4895 WorkList ID : 144502 Department : Wet-Chemistry Date : 11-25-2020 13:54:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/25/2020	Water	L4895-01	pH	Cool 4 deg C	PSEG03	J12	COMP-56-67	11/25/2020	9040C
11/25/2020	Water	L4895-03	pH	Cool 4 deg C	PSEG03	J12	NWB- 1558	11/25/2020	9040C
11/25/2020	Water	L4896-01	pH	Cool 4 deg C	PSEG03	J12	1861	11/25/2020	9040C
11/25/2020	Water	L4896-02	pH	Cool 4 deg C	PSEG03	J12	1862	11/25/2020	9040C
11/25/2020	Water	L4896-03	pH	Cool 4 deg C	PSEG03	J12	1863	11/25/2020	9040C
11/25/2020	Water	L4896-04	pH	Cool 4 deg C	PSEG03	J12	1864	11/25/2020	9040C
11/25/2020	Water	L4896-05	pH	Cool 4 deg C	PSEG03	J12	1865	11/25/2020	9040C
11/25/2020	Water	L4896-06	pH	Cool 4 deg C	PSEG03	J12	1866	11/25/2020	9040C
11/25/2020	Water	L4896-07	pH	Cool 4 deg C	PSEG03	J12	1867	11/25/2020	9040C
11/25/2020	Water	L4896-08	pH	Cool 4 deg C	PSEG03	J12	1868	11/25/2020	9040C
11/25/2020	Water	L4896-09	pH	Cool 4 deg C	PSEG03	J12	1869	11/25/2020	9040C
11/25/2020	Water	L4896-10	pH	Cool 4 deg C	PSEG03	J12	1870	11/25/2020	9040C

Date/Time 11-25-2020 19:00
 Received by: JP Westberg
 Relinquished by: JP

Date/Time 11-25-2020 15:35
 Received by: JP
 Relinquished by: JP Westberg