



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112144

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/23/2020	16:40
2	CAL2	1	Water	NA	NA	20.2	7.00	11/23/2020	16:41
3	CAL3	1	Water	NA	NA	20.2	10.03	11/23/2020	16:42
4	ICV	1	Water	NA	NA	20.2	6.99	11/23/2020	16:43
5	CCV1	1	Water	NA	NA	20.2	2.01	11/23/2020	16:45
6	L4597-01	1	Solid	20.02	20	21.9	7.21	11/23/2020	16:48
7	L4597-01DUP	1	Solid	20.03	20	21.8	7.22	11/23/2020	16:49
8	CCV2	1	Water	NA	NA	20.2	12.02	11/23/2020	16:50

187 112144

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4597 WorkList ID : 144503 Department : Wet-Chemistry Date : 11-23-2020 16:30:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/23/2020	Solid	L4597-01	pH	Cool 4 deg C	PSEG05	H53	BU-03-112320	11/23/2020	9045D

Date/Time 11/23/2020 16:35
Received by: JD (w/degm)
Relinquished by: JD

Date/Time 11/23/2020 18:00
Received by: JD
Relinquished by: JD (w/degm) Creng