



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112147

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/25/2020	16:30
2	CAL2	1	Water	NA	NA	20.1	7.01	11/25/2020	16:31
3	CAL3	1	Water	NA	NA	20.1	10.02	11/25/2020	16:32
4	ICV	1	Water	NA	NA	20.2	6.99	11/25/2020	16:33
5	CCV1	1	Water	NA	NA	20.3	2.01	11/25/2020	16:35
6	L4899-04	1	Solid	20.03	20	21.4	8.52	11/25/2020	16:36
7	L4900-04	1	Solid	20.04	20	21.8	10.86	11/25/2020	16:37
8	L4900-06	1	Solid	20.02	20	21.8	10.03	11/25/2020	16:38
9	L4903-04	1	Solid	20.01	20	21.3	9.39	11/25/2020	16:40
10	L4903-08	1	Solid	20.03	20	21.3	8.37	11/25/2020	16:42
11	L4903-08DUP	1	Solid	20.02	20	21.2	8.38	11/25/2020	16:43
12	CCV2	1	Water	NA	NA	20.1	12.02	11/25/2020	16:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity I4899

WorkList ID : 144497

Department : Wet-Chemistry

Date : 11-25-2020 13:00:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/25/2020	Solid	L4899-04	Corrosivity	Cool 4 deg C	PSEG03	J12	TP-1-2	11/25/2020	9045D
11/25/2020	Solid	L4900-04	Corrosivity	Cool 4 deg C	ENTE01	J23	PAV-B2-112520-0-10	11/25/2020	9045D
11/25/2020	Solid	L4900-06	Corrosivity	Cool 4 deg C	ENTE01	J23	PAV-B2-112520-10-12.75	11/25/2020	9045D
11/25/2020	Solid	L4903-04	Corrosivity	Cool 4 deg C	PSEG03	J12	TP-1-2-COMP	11/25/2020	9045D
11/25/2020	Solid	L4903-08	Corrosivity	Cool 4 deg C	PSEG03	J12	TP-3-4-COMP	11/25/2020	9045D

Date/Time 11-25-2020 13:10

Received by: JP (wetchem)

Relinquished by: KS (wetchem)

Date/Time 11-25-2020 17:00

Received by: KS (wetchem)

Relinquished by: JP (wetchem)