

LB112153


7470A

INSTRUMENT ID: CVI

Linear

 μ Abs.:

35251



A= 0.0000e+000

B= 2.8630e-004

C= 1.1531e-002

Rho= 0.9998060

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	4	0.000	4				
0.2	0.200	0.212	0.012	699	0.0 %	699				
2.5	2.500	2.548	0.048	8861	0.0 %	8861				
5.0	5.000	4.948	-0.052	17241	0.0 %	17241				
7.5	7.500	7.376	-0.124	25721	0.0 %	25721				
10.0	10.000	10.104	0.104	35251	0.0 %	35251				

LB112153 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Nov 2020 14:02:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000	7470A	PPB	27 Nov 2020 14:08:35	S	Std
0.2 - 1	50.2	699	-	0.2000	7470A	PPB	27 Nov 2020 14:10:52	S	Std
2.5 - 1	52.5	8861	-	2.5000	7470A	PPB	27 Nov 2020 14:13:09	S	Std
5.0 - 1	55	17241	-	5.0000	7470A	PPB	27 Nov 2020 14:15:26	S	Std
7.5 - 1	57.5	25721	-	7.5000	7470A	PPB	27 Nov 2020 14:20:22	S	Std
10.0 - 1	510	35251	-	10.0000	7470A	PPB	27 Nov 2020 14:30:19	S	Std
ICV19 - 1	ICV19	14129	4.0567	-	7470A	PPB	27 Nov 2020 14:33:25	U	SMPL
ICB19 - 1	ICB19	-83	-0.0122	-	7470A	PPB	27 Nov 2020 14:35:40	U	SMPL
CCV01 - 1	CCV01	17358	4.9812	-	7470A	PPB	27 Nov 2020 14:37:55	U	SMPL
CCB01 - 1	CCB01	-99	-0.0168	-	7470A	PPB	27 Nov 2020 14:40:11	U	SMPL
CRA - 1	CRA	669	0.2031	-	7470A	PPB	27 Nov 2020 14:42:26	U	SMPL
HighStd - 1	HighStd	35200	10.0894	-	7470A	PPB	27 Nov 2020 14:44:41	U	SMPL
ChkStd - 1	ChkStd	24008	6.8851	-	7470A	PPB	27 Nov 2020 14:46:56	U	SMPL
PB133289BL - 1	PBW	-78	-0.0108	-	7470A	PPB	27 Nov 2020 14:49:13	U	SMPL
PB133289BS - 1	LCSW	15580	4.4721	-	7470A	PPB	27 Nov 2020 14:51:29	U	SMPL
L4895-01 - 1	COMP-56-67	-93	-0.0151	-	7470A	PPB	27 Nov 2020 14:53:47	U	SMPL
L4896-11 - 1	61-70	-16	0.0070	-	7470A	PPB	27 Nov 2020 14:56:03	U	SMPL
L4896-11DUP - 1	61-70DUP	8	0.0138	-	7470A	PPB	27 Nov 2020 14:58:20	U	SMPL
L4896-11MS - 1	61-70MS	14277	4.0991	-	7470A	PPB	27 Nov 2020 15:00:38	U	SMPL
L4896-11MSD - 1	61-70MSD	14165	4.0670	-	7470A	PPB	27 Nov 2020 15:02:53	U	SMPL
CCV02 - 1	CCV02	17416	4.9978	-	7470A	PPB	27 Nov 2020 15:05:08	U	SMPL
CCB02 - 1	CCB02	-94	-0.0154	-	7470A	PPB	27 Nov 2020 15:07:24	U	SMPL
PB133289TB - 1	PB133267TB	-4	0.0104	-	7470A	PPB	27 Nov 2020 15:09:39	U	SMPL
L4896-11LX5 - 1		-7	0.0095	-	7470A	PPB	27 Nov 2020 15:11:54	U	SMPL
L4896-11A - 1		14248	4.0908	-	7470A	PPB	27 Nov 2020 15:14:10	U	SMPL
CCV03 - 1	CCV03	17098	4.9067	-	7470A	PPB	27 Nov 2020 15:16:25	U	SMPL
CCB03 - 1	CCB03	-93	-0.0151	-	7470A	PPB	27 Nov 2020 15:18:41	U	SMPL