

LB112155


7471B

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

49542



A= 0.0000e+000

B= 1.9818e-004

C= 1.0468e-002

Rho= 0.9993919

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	39	0.000	39				
0.2	0.200	0.206	0.006	987	0.0 %	987				
2.5	2.500	2.383	-0.117	11974	0.0 %	11974				
5.0	5.000	5.029	0.029	25324	0.0 %	25324				
7.5	7.500	7.735	0.235	38977	0.0 %	38977				
10.0	10.000	9.828	-0.172	49542	0.0 %	49542				

LB112155 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Nov 2020 16:22:10

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	39	-	0.00007471B	PPB	27 Nov 2020 16:27:21	S	Std
0.2 - 1	50.2	987	-	0.20007471B	PPB	27 Nov 2020 16:29:38	S	Std
2.5 - 1	52.5	11974	-	2.50007471B	PPB	27 Nov 2020 16:31:55	S	Std
5.0 - 1	55	25324	-	5.00007471B	PPB	27 Nov 2020 16:34:12	S	Std
7.5 - 1	57.5	38977	-	7.50007471B	PPB	27 Nov 2020 16:36:29	S	Std
10.0 - 1	510	49542	-	10.00007471B	PPB	27 Nov 2020 16:41:23	S	Std
ICV20 - 1	ICV20	19445	3.8640	-7471B	PPB	27 Nov 2020 16:44:02	U	SMPL
ICB20 - 1	ICB20	-84	-0.0062	-7471B	PPB	27 Nov 2020 16:46:17	U	SMPL
CCV04 - 1	CCV04	25695	5.1026	-7471B	PPB	27 Nov 2020 16:48:32	U	SMPL
CCB04 - 1	CCB04	-107	-0.0107	-7471B	PPB	27 Nov 2020 16:50:48	U	SMPL
CRA - 1	CRA	979	0.2045	-7471B	PPB	27 Nov 2020 16:53:03	U	SMPL
HighStd - 1	HighStd	49933	9.9060	-7471B	PPB	27 Nov 2020 16:55:18	U	SMPL
ChkStd - 1	ChkStd	37050	7.3529	-7471B	PPB	27 Nov 2020 16:57:33	U	SMPL
PB133293BL - 1	PBS	-144	-0.0181	-7471B	PPB	27 Nov 2020 16:59:50	U	SMPL
PB133293BS - 1	LCSS	20101	3.9940	-7471B	PPB	27 Nov 2020 17:02:07	U	SMPL
L4899-03 - 1	TP-1-2	4613	0.9247	-7471B	PPB	27 Nov 2020 17:04:23	U	SMPL
L4900-03 - 1	PAV-B2-112520-0-10	22297	4.4292	-7471B	PPB	27 Nov 2020 17:06:39	U	SMPL
L4900-05 - 1	PAV-B2-112520-10-12.75	10404	2.0723	-7471B	PPB	27 Nov 2020 17:13:45	U	SMPL
L4900-05DUP - 1	PAV-B2-112520-10-12.75DUP	9565	1.9060	-7471B	PPB	27 Nov 2020 17:16:03	U	SMPL
L4900-05MS - 1	PAV-B2-112520-10-12.75MS	29292	5.8154	-7471B	PPB	27 Nov 2020 17:20:55	U	SMPL
CCV05 - 1	CCV05	25055	4.9758	-7471B	PPB	27 Nov 2020 17:23:10	U	SMPL
CCB05 - 1	CCB05	-81	-0.0056	-7471B	PPB	27 Nov 2020 17:25:26	U	SMPL
L4900-05MSD - 1	PAV-B2-112520-10-12.75MSD	30797	6.1137	-7471B	PPB	27 Nov 2020 17:27:41	U	SMPL
L4903-03 - 1	TP-1-2-COMP	1795	0.3662	-7471B	PPB	27 Nov 2020 17:29:56	U	SMPL
L4903-07 - 1	TP-3-4-COMP	3900	0.7834	-7471B	PPB	27 Nov 2020 17:32:12	U	SMPL
L4900-05LX5 - 1		2026	0.4120	-7471B	PPB	27 Nov 2020 17:34:28	U	SMPL
L4900-05A - 1		27614	5.4829	-7471B	PPB	27 Nov 2020 17:36:44	U	SMPL
CCV06 - 1	CCV06	23934	4.7536	-7471B	PPB	27 Nov 2020 17:39:00	U	SMPL
CCB06 - 1	CCB06	-77	-0.0048	-7471B	PPB	27 Nov 2020 17:41:17	U	SMPL