

LB112164


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35033



A= 0.0000e+000

B= 2.8510e-004

C= 2.5062e-002

Rho= 0.9999589

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	-10	0.000	-10				
0.2	0.200	0.205	0.005	630	0.0 %	630				
2.5	2.500	2.504	0.004	8694	0.0 %	8694				
5.0	5.000	4.928	-0.072	17197	0.0 %	17197				
7.5	7.500	7.529	0.029	26319	0.0 %	26319				
10.0	10.000	10.013	0.013	35033	0.0 %	35033				

LB112164 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 30 Nov 2020 12:26:25							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-10	-	0.00007470A	PPB		30 Nov 2020 13:28:55	S	Std
0.2 - 1	50.2	630	-	0.20007470A	PPB		30 Nov 2020 13:31:12	S	Std
2.5 - 1	52.5	8694	-	2.50007470A	PPB		30 Nov 2020 13:33:29	S	Std
5.0 - 1	55	17197	-	5.00007470A	PPB		30 Nov 2020 13:35:46	S	Std
7.5 - 1	57.5	26319	-	7.50007470A	PPB		30 Nov 2020 13:38:03	S	Std
10.0 - 1	510	35033	-	10.00007470A	PPB		30 Nov 2020 13:40:19	S	Std
ICV21 - 1	ICV21	13057	3.7476	-7470A	PPB		30 Nov 2020 13:45:30	U	SMPL
ICB21 - 1	ICB21	-62	0.0074	-7470A	PPB		30 Nov 2020 13:47:45	U	SMPL
CCV07 - 1	CCV07	16474	4.7218	-7470A	PPB		30 Nov 2020 13:50:01	U	SMPL
CCB07 - 1	CCB07	-93	-0.0015	-7470A	PPB		30 Nov 2020 13:52:16	U	SMPL
CRA - 1	CRA	581	0.1907	-7470A	PPB		30 Nov 2020 13:54:32	U	SMPL
HighStd - 1	HighStd	34023	9.7250	-7470A	PPB		30 Nov 2020 13:56:47	U	SMPL
ChkStd - 1	ChkStd	22591	6.4657	-7470A	PPB		30 Nov 2020 14:06:07	U	SMPL
PB133308BL - 1	PBW	-121	-0.0094	-7470A	PPB		30 Nov 2020 14:08:22	U	SMPL
PB133308BS - 1	LCSW	11377	3.2686	-7470A	PPB		30 Nov 2020 14:10:38	U	SMPL
L4706-04 - 1	16874	323	0.1171	-7470A	PPB		30 Nov 2020 14:12:55	U	SMPL
L4891-04 - 1	PAV-B3-112420-0-10	38	0.0359	-7470A	PPB		30 Nov 2020 14:17:00	U	SMPL
L4899-04 - 1	TP-1-2	-26	0.0176	-7470A	PPB		30 Nov 2020 14:21:34	U	SMPL
L4900-04 - 1	PAV-B2-112520-0-10	3	0.0259	-7470A	PPB		30 Nov 2020 14:25:17	U	SMPL
CCV08 - 1	CCV08	18143	5.1976	-7470A	PPB		30 Nov 2020 14:27:32	U	SMPL
CCB08 - 1	CCB08	-80	0.0023	-7470A	PPB		30 Nov 2020 14:29:47	U	SMPL
L4900-06 - 1	PAV-B2-112520-10-12.75	5	0.0265	-7470A	PPB		30 Nov 2020 14:32:45	U	SMPL
L4903-04 - 1	TP-1-2-COMP	-5	0.0236	-7470A	PPB		30 Nov 2020 14:35:33	U	SMPL
L4903-08 - 1	TP-3-4-COMP	1	0.0253	-7470A	PPB		30 Nov 2020 14:37:49	U	SMPL
L4903-08DUP - 1	TP-3-4-COMPDUP	-12	0.0216	-7470A	PPB		30 Nov 2020 14:40:04	U	SMPL
L4903-08MS - 1	TP-3-4-COMPMS	14650	4.2018	-7470A	PPB		30 Nov 2020 14:42:20	U	SMPL
L4903-08MSD - 1	TP-3-4-COMPMSD	14346	4.1151	-7470A	PPB		30 Nov 2020 14:44:37	U	SMPL
PB133308TB - 1	PB133265TB	-43	0.0128	-7470A	PPB		30 Nov 2020 14:48:48	U	SMPL
L4903-08LX5 - 1		0	0.0251	-7470A	PPB		30 Nov 2020 14:51:05	U	SMPL
L4903-08A - 1		14514	4.1630	-7470A	PPB		30 Nov 2020 14:53:22	U	SMPL
L4891-06 - 1	PAV-B3-112420-10-11.5	-12	0.0216	-7470A	PPB		30 Nov 2020 14:55:37	U	SMPL
CCV09 - 1	CCV09	18619	5.3333	-7470A	PPB		30 Nov 2020 14:57:53	U	SMPL
CCB09 - 1	CCB09	-71	0.0048	-7470A	PPB		30 Nov 2020 15:00:08	U	SMPL