



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112170

Ambient Barometric Pressure (mmHg): 745.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	81.5	11/30/2020	11:35
2	L4895-01		1	8	100.00	>212.0	>212.0	11/30/2020	12:10
3	L4895-01DUP		1	8	100.00	>212.0	>212.0	11/30/2020	12:40
4	L4895-03		1	9	100.00	>212.0	>212.0	11/30/2020	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp113020 WorkList ID : 144545 Department : Wet-Chemistry Date : 11-30-2020 09:07:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/08/2020	Water	L4895-01	Flash Point	Cool 4 deg C	PSEG03	J12	COMP-56-67	11/25/2020	1010A
12/08/2020	Water	L4895-03	Flash Point	Cool 4 deg C	PSEG03	J12	NWB- 1558	11/25/2020	1010A

Date/Time 11/30/20 11:00
 Received by: KS (Wetchem)
 Relinquished by: SP (Wetchem)

Date/Time 11/30/20 13:20
 Received by: SP (Wetchem)
 Relinquished by: KS (Wetchem)