



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112174

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/30/2020	16:48
2	CAL2	1	Water	NA	NA	20.1	7.01	11/30/2020	16:49
3	CAL3	1	Water	NA	NA	20.1	10.03	11/30/2020	16:50
4	ICV	1	Water	NA	NA	20.1	6.98	11/30/2020	16:51
5	CCV1	1	Water	NA	NA	20.2	2.01	11/30/2020	16:52
6	L4910-04	1	Solid	20.03	20	22.3	10.12	11/30/2020	17:00
7	L4910-04DUP	1	Solid	20.04	20	22.4	10.13	11/30/2020	17:01
8	L4910-06	1	Solid	20.02	20	22.2	11.76	11/30/2020	17:03
9	CCV2	1	Water	NA	NA	20.1	12.02	11/30/2020	17:05

WORKLIST(Hardcopy Internal Chain)

YB 11/21/20

WorkList Name : corrosivity I4906

WorkList ID : 144557

Department : Wet-Chemistry

Date : 11-30-2020 16:12:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/30/2020	Solid	L4910-04	Corrosivity	Cool 4 deg C	ENTE01	J31	PAV-B47-113020-0-10	11/30/2020	9045D
11/30/2020	Solid	L4910-06	Corrosivity	Cool 4 deg C	ENTE01	J31	PAV-B47-113020-10-12.75	11/30/2020	9045D

Date/Time 11/30/2020 16:15
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)

Date/Time 11/30/2020 17:35
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)