



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112176

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/30/2020	16:48
2	CAL2	1	Water	NA	NA	20.1	7.01	11/30/2020	16:49
3	CAL3	1	Water	NA	NA	20.1	10.03	11/30/2020	16:50
4	ICV	1	Water	NA	NA	20.1	6.98	11/30/2020	16:51
5	CCV1	1	Water	NA	NA	20.2	2.01	11/30/2020	16:52
6	L4905-06	1	Water	NA	NA	17.2	5.66	11/30/2020	17:15
7	L4905-08	1	Water	NA	NA	17.2	5.59	11/30/2020	17:18
8	L4905-10	1	Water	NA	NA	17.3	10.16	11/30/2020	17:20
9	L4905-11	1	Water	NA	NA	17.7	5.78	11/30/2020	17:21
10	L4905-12	1	Water	NA	NA	17.2	5.97	11/30/2020	17:23
11	L4905-13	1	Water	NA	NA	17.2	5.15	11/30/2020	17:25
12	L4905-14	1	Water	NA	NA	17.3	5.51	11/30/2020	17:26
13	L4905-15	1	Water	NA	NA	17.2	5.88	11/30/2020	17:29
14	L4905-16	1	Water	NA	NA	17.5	6.07	11/30/2020	17:30
15	L4905-17	1	Water	NA	NA	17.6	5.21	11/30/2020	17:32
16	CCV2	1	Water	NA	NA	20.1	12.02	11/30/2020	17:35
17	L4905-18	1	Water	NA	NA	17.6	5.75	11/30/2020	17:36
18	L4905-19	1	Water	NA	NA	18.0	5.65	11/30/2020	17:38
19	L4905-19DUP	1	Water	NA	NA	18.1	5.66	11/30/2020	17:39
20	CCV3	1	Water	NA	NA	20.2	2.01	11/30/2020	17:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH L4905 W WorkList ID: 144558 Department: Wet-Chemistry Date: 11-30-2020 16:13:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/30/2020	Water	L4905-06	pH	Cool 4 deg C	PSEG03	J21	COMP-06-07-08-09-10	11/30/2020	9040C
11/30/2020	Water	L4905-08	pH	Cool 4 deg C	PSEG03	J21	COMP-11-12-13-14	11/30/2020	9040C
11/30/2020	Water	L4905-10	pH	Cool 4 deg C	PSEG03	J21	32778	11/30/2020	9040C
11/30/2020	Water	L4905-11	pH	Cool 4 deg C	PSEG03	J21	32806	11/30/2020	9040C
11/30/2020	Water	L4905-12	pH	Cool 4 deg C	PSEG03	J21	32807	11/30/2020	9040C
11/30/2020	Water	L4905-13	pH	Cool 4 deg C	PSEG03	J21	32808	11/30/2020	9040C
11/30/2020	Water	L4905-14	pH	Cool 4 deg C	PSEG03	J21	32809	11/30/2020	9040C
11/30/2020	Water	L4905-15	pH	Cool 4 deg C	PSEG03	J21	32810	11/30/2020	9040C
11/30/2020	Water	L4905-16	pH	Cool 4 deg C	PSEG03	J21	32811	11/30/2020	9040C
11/30/2020	Water	L4905-17	pH	Cool 4 deg C	PSEG03	J21	32812	11/30/2020	9040C
11/30/2020	Water	L4905-18	pH	Cool 4 deg C	PSEG03	J21	32813	11/30/2020	9040C
11/30/2020	Water	L4905-19	pH	Cool 4 deg C	PSEG03	J21	32814	11/30/2020	9040C

Date/Time 11-30-2020 16:15
 Received by: JS (wet chem)
 Relinquished by: JS (wet chem)

Date/Time 11-30-2020 17:25
 Received by: JS (wet chem)
 Relinquished by: JS (wet chem)