

LB 112180


INSTRUMENT ID: CV1

7470A

Linear

 μ Abs.:

37720



A= 0.0000e+000

B= 2.6648e-004

C= 1.8643e-002

Rho= 0.9998488

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	18	0.000	18				
0.2	0.200	0.200	0.000	681	0.0 %	681				
2.5	2.500	2.478	-0.022	9228	0.0 %	9228				
5.0	5.000	5.054	0.054	18896	0.0 %	18896				
7.5	7.500	7.374	-0.126	27603	0.0 %	27603				
10.0	10.000	10.070	0.070	37720	0.0 %	37720				

LB112180 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 01 Dec 2020 13:03:17			
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type
0.0 - 1	50	18	-	0.00007470A	PPB	01 Dec 2020 13:33:10	S Std
0.2 - 1	50.2	681	-	0.20007470A	PPB	01 Dec 2020 13:35:27	S Std
2.5 - 1	52.5	9228	-	2.50007470A	PPB	01 Dec 2020 13:37:44	S Std
5.0 - 1	55	18896	-	5.00007470A	PPB	01 Dec 2020 13:40:01	S Std
7.5 - 1	57.5	27603	-	7.50007470A	PPB	01 Dec 2020 13:42:17	S Std
10.0 - 1	511	37720	-	10.00007470A	PPB	01 Dec 2020 13:44:33	S Std
ICV22 - 1	ICV22	13688	3.6662	-7470A	PPB	01 Dec 2020 13:47:13	U SMPL
ICB22 - 1	ICB22	-74	-0.0011	-7470A	PPB	01 Dec 2020 13:49:28	U SMPL
CCV10 - 1	CCV10	18966	5.0727	-7470A	PPB	01 Dec 2020 13:51:43	U SMPL
CCB10 - 1	CCB10	-60	0.0027	-7470A	PPB	01 Dec 2020 13:53:58	U SMPL
CRA - 1	CRA	668	0.1967	-7470A	PPB	01 Dec 2020 13:56:14	U SMPL
HighStd - 1	HighStd	38277	10.2187	-7470A	PPB	01 Dec 2020 13:58:29	U SMPL
ChkStd - 1	ChkStd	24546	6.5597	-7470A	PPB	01 Dec 2020 14:00:44	U SMPL
PB133329BL - 1	PBW	-123	-0.0141	-7470A	PPB	01 Dec 2020 14:03:01	U SMPL
PB133329BS - 1	LCSW	14005	3.7507	-7470A	PPB	01 Dec 2020 14:05:16	U SMPL
L4592-05 - 1	EO-01-112020	-80	-0.0027	-7470A	PPB	01 Dec 2020 14:07:33	U SMPL
L4905-01 - 1	COMP-00-01-02	-25	0.0120	-7470A	PPB	01 Dec 2020 14:09:49	U SMPL
L4905-03 - 1	COMP-03-04-05	19	0.0237	-7470A	PPB	01 Dec 2020 14:12:06	U SMPL
L4910-04 - 1	PAV-B47-113020-0-10	40	0.0293	-7470A	PPB	01 Dec 2020 14:14:23	U SMPL
L4910-06 - 1	PAV-B47-113020-10-12.75	-10	0.0160	-7470A	PPB	01 Dec 2020 14:16:38	U SMPL
CCV11 - 1	CCV11	18671	4.9941	-7470A	PPB	01 Dec 2020 14:18:54	U SMPL
CCB11 - 1	CCB11	-51	0.0051	-7470A	PPB	01 Dec 2020 14:21:10	U SMPL
L4910-06DUP - 1	PAV-B47-113020-10-12.75DUP	17	0.0232	-7470A	PPB	01 Dec 2020 14:23:25	U SMPL
L4910-06MS - 1	PAV-B47-113020-10-12.75MS	14066	3.7670	-7470A	PPB	01 Dec 2020 14:25:40	U SMPL
L4910-06MSD - 1	PAV-B47-113020-10-12.75MSD	14208	3.8048	-7470A	PPB	01 Dec 2020 14:27:56	U SMPL
PB133329TB - 1	PB133319TB	-76	-0.0016	-7470A	PPB	01 Dec 2020 14:30:12	U SMPL
L4910-06LX5 - 1		-4	0.0176	-7470A	PPB	01 Dec 2020 14:32:28	U SMPL
L4910-06A - 1		14956	4.0041	-7470A	PPB	01 Dec 2020 14:34:45	U SMPL
CCV12 - 1	CCV12	18488	4.9453	-7470A	PPB	01 Dec 2020 14:37:01	U SMPL
CCB12 - 1	CCB12	-75	-0.0013	-7470A	PPB	01 Dec 2020 14:39:18	U SMPL