


LB112182
INSTRUMENT ID: CV1

7471B

Linear



A= 0.0000e+000

B= 1.8912e-004

C= 2.6753e-003

Rho= 0.9995030

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	17	0.000	17				
0.2	0.200	0.195	-0.005	1016	0.0 %	1016				
2.5	2.500	2.446	-0.054	12921	0.0 %	12921				
5.0	5.000	4.980	-0.020	26319	0.0 %	26319				
7.5	7.500	7.730	0.230	40862	0.0 %	40862				
10.0	10.000	9.843	-0.157	52031	0.0 %	52031				

LB112182 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 01 Dec 2020 14:48:22				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	17	-	0.00007471B	PPB	01 Dec 2020 14:49:21	S	Std
0.2 - 1	50.2	1016	-	0.20007471B	PPB	01 Dec 2020 14:51:38	S	Std
2.5 - 1	52.5	12921	-	2.50007471B	PPB	01 Dec 2020 14:53:56	S	Std
5.0 - 1	55	26319	-	5.00007471B	PPB	01 Dec 2020 14:56:13	S	Std
7.5 - 1	57.5	40862	-	7.50007471B	PPB	01 Dec 2020 14:58:30	S	Std
10.0 - 1	59	52031	-	10.00007471B	PPB	01 Dec 2020 15:00:46	S	Std
ICV23 - 1	ICV23	20132	3.8100	-7471B	PPB	01 Dec 2020 15:03:28	U	SMPL
ICB23 - 1	ICB23	-84	-0.0132	-7471B	PPB	01 Dec 2020 15:05:43	U	SMPL
CCV13 - 1	CCV13	26352	4.9863	-7471B	PPB	01 Dec 2020 15:08:00	U	SMPL
CCB13 - 1	CCB13	-96	-0.0155	-7471B	PPB	01 Dec 2020 15:10:16	U	SMPL
CRA - 1	CRA	988	0.1895	-7471B	PPB	01 Dec 2020 15:12:31	U	SMPL
HighStd - 1	HighStd	52779	9.9841	-7471B	PPB	01 Dec 2020 15:14:47	U	SMPL
ChkStd - 1	ChkStd	34333	6.4956	-7471B	PPB	01 Dec 2020 15:17:02	U	SMPL
PB133328BL - 1	PBS	-120	-0.0200	-7471B	PPB	01 Dec 2020 15:19:19	U	SMPL
PB133328BS - 1	LCSS	18458	3.4934	-7471B	PPB	01 Dec 2020 15:21:36	U	SMPL
L4906-01 - 1	4-DRUM	1512	0.2886	-7471B	PPB	01 Dec 2020 15:28:54	U	SMPL
L4906-01DUP - 1	4-DRUMDUP	1437	0.2744	-7471B	PPB	01 Dec 2020 15:31:10	U	SMPL
L4906-01MS - 1	4-DRUMMS	23894	4.5214	-7471B	PPB	01 Dec 2020 15:33:27	U	SMPL
L4906-01MSD - 1	4-DRUMMSD	22568	4.2707	-7471B	PPB	01 Dec 2020 15:35:44	U	SMPL
L4909-01 - 1	ROCK-HILL-SLOPE-EMBANKMENT-A	806	0.1551	-7471B	PPB	01 Dec 2020 15:37:59	U	SMPL
CCV14 - 1	CCV14	26504	5.0150	-7471B	PPB	01 Dec 2020 15:40:15	U	SMPL
CCB14 - 1	CCB14	-83	-0.0130	-7471B	PPB	01 Dec 2020 15:42:30	U	SMPL
L4909-02 - 1	ROCK-HILL-SLOPE-EMBANKMENT-B	3182	0.6044	-7471B	PPB	01 Dec 2020 15:44:46	U	SMPL
L4910-03 - 1	PAV-B47-113020-0-10	9674	1.8322	-7471B	PPB	01 Dec 2020 15:47:01	U	SMPL
L4910-05 - 1	PAV-B47-113020-10-12.75	12642	2.3935	-7471B	PPB	01 Dec 2020 15:49:17	U	SMPL
L4906-01LX5 - 1		220	0.0443	-7471B	PPB	01 Dec 2020 15:51:32	U	SMPL
L4906-01A - 1		23047	4.3613	-7471B	PPB	01 Dec 2020 15:53:49	U	SMPL
CCV15 - 1	CCV15	26614	5.0358	-7471B	PPB	01 Dec 2020 15:56:05	U	SMPL
CCB15 - 1	CCB15	-72	-0.0109	-7471B	PPB	01 Dec 2020 15:58:21	U	SMPL