



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112197

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	12/02/2020	14:15
2	CAL2	1	Water	NA	NA	21.2	7.00	12/02/2020	14:16
3	CAL3	1	Water	NA	NA	21.2	10.03	12/02/2020	14:18
4	ICV	1	Water	NA	NA	21.2	6.98	12/02/2020	14:20
5	CCV1	1	Water	NA	NA	21.2	2.01	12/02/2020	14:22
6	L4918-02	1	Solid	20.02	20	21.6	6.36	12/02/2020	14:25
7	L4918-02DUP	1	Solid	20.01	20	21.7	6.37	12/02/2020	14:26
8	L4918-04	1	Solid	20.03	20	21.7	5.78	12/02/2020	14:27
9	L4918-06	1	Solid	20.05	20	20.9	4.70	12/02/2020	14:30
10	CCV2	1	Water	NA	NA	21.2	12.02	12/02/2020	14:31

13112197

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4918 WorkList ID : 144606 Department : Wet-Chemistry Date : 12-02-2020 14:03:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/02/2020	Solid	L4918-02	pH	Cool 4 deg C	PSEG03	J23	20801	12/02/2020	9045D
12/02/2020	Solid	L4918-04	pH	Cool 4 deg C	PSEG03	J23	16827	12/02/2020	9045D
12/02/2020	Solid	L4918-06	pH	Cool 4 deg C	PSEG03	J23	16829	12/02/2020	9045D

Date/Time 12-02-2020 14:10
Received by: JP (cust - Open)
Relinquished by: AP (cust - Open)

Date/Time 12-02-2020 15:30
Received by: 70 (cust - Open)
Relinquished by: 70 (cust - Open)