



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112201

Ambient Barometric Pressure (mmHg): 765.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	12/03/2020	11:35
2	L4905-06		1	8	100.00	>212.0	>212.0	12/03/2020	12:10
3	L4905-06DUP		1	8	100.00	>212.0	>212.0	12/03/2020	12:40
4	L4905-08		1	9	100.00	>212.0	>212.0	12/03/2020	13:15
5	L4905-10		1	9	100.00	>212.0	>212.0	12/03/2020	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp12320 WorkList ID : 144625 Department : Wet-Chemistry Date : 12-03-2020 09:02:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2020	Water	L4905-06	Flash Point	Cool 4 deg C	PSEG03	J21	COMP-06-07-08-09-10	11/30/2020	1010A
12/07/2020	Water	L4905-08	Flash Point	Cool 4 deg C	PSEG03	J21	COMP-11-12-13-14	11/30/2020	1010A
12/07/2020	Water	L4905-10	Flash Point	Cool 4 deg C	PSEG03	J21	32778	11/30/2020	1010A

Date/Time 12/03/2020 11:10
 Received by: KS (Wetchem)
 Relinquished by: fp12320 (Wetchem)

Date/Time 12/03/2020 13:50
 Received by: AP (Ext. Lab)
 Relinquished by: KS (Wetchem)