



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112205

Slope : 97.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	12/03/2020	12:20
2	CAL2	1	Water	NA	NA	21.1	7.00	12/03/2020	12:21
3	CAL3	1	Water	NA	NA	21.1	10.03	12/03/2020	12:22
4	ICV	1	Water	NA	NA	21.2	6.98	12/03/2020	12:24
5	CCV1	1	Water	NA	NA	21.1	2.01	12/03/2020	12:25
6	L4934-01	1	Water	NA	NA	13.2	6.63	12/03/2020	12:26
7	L4934-01DUP	1	Water	NA	NA	13.3	6.64	12/03/2020	12:27
8	CCV2	1	Water	NA	NA	21.1	12.02	12/03/2020	12:28

112205

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14934 WorkList ID : 144638 Department : Wet-Chemistry Date : 12-03-2020 12:14:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/02/2020	Water	L4934-01	pH	Cool 4 deg C	APT101	J11	WGGM-20-WW-TSF-041	12/02/2020	SM 4500-H B

Date/Time 12-03-2020 12:15
 Received by: JL West grom
 Relinquished by: [Signature] SW

Date/Time 12-03-2020 13:00
 Received by: [Signature] SW
 Relinquished by: SA West Green