

LB112207


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

54531



A= 0.0000e+000

B= 1.8494e-004

C= -2.7930e-003

Rho= 0.9998798

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	322	0.000	322				
0.2	0.200	0.212	0.012	1164	0.0 %	1164				
2.5	2.500	2.481	-0.019	13433	0.0 %	13433				
5.0	5.000	4.920	-0.080	26620	0.0 %	26620				
7.5	7.500	7.447	-0.053	40282	0.0 %	40282				
10.0	10.000	10.082	0.082	54531	0.0 %	54531				

LB112207 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Dec 2020 13:24:19

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	322	-	0.00007471B	PPB	03 Dec 2020 13:31:20	S	Std
0.2 - 1	50.2	1164	-	0.20007471B	PPB	03 Dec 2020 13:33:37	S	Std
2.5 - 1	52.5	13433	-	2.50007471B	PPB	03 Dec 2020 13:35:54	S	Std
5.0 - 1	55	26620	-	5.00007471B	PPB	03 Dec 2020 13:38:11	S	Std
7.5 - 1	57.5	40282	-	7.50007471B	PPB	03 Dec 2020 13:40:28	S	Std
10.0 - 1	57.6	54531	-	10.00007471B	PPB	03 Dec 2020 13:42:44	S	Std
ICV26 - 1	ICV26	22052	4.0755	-7471B	PPB	03 Dec 2020 13:46:25	U	SMPL
ICB26 - 1	ICB26	54	0.0072	-7471B	PPB	03 Dec 2020 13:50:55	U	SMPL
CCV22 - 1	CCV22	27008	4.9920	-7471B	PPB	03 Dec 2020 13:53:10	U	SMPL
CCB22 - 1	CCB22	34	0.0035	-7471B	PPB	03 Dec 2020 13:57:54	U	SMPL
CRA - 1	CRA	1197	0.2186	-7471B	PPB	03 Dec 2020 14:00:09	U	SMPL
HighStd - 1	HighStd	55397	10.2423	-7471B	PPB	03 Dec 2020 14:02:24	U	SMPL
ChkStd - 1	ChkStd	35281	6.5220	-7471B	PPB	03 Dec 2020 14:04:40	U	SMPL
PB133365BL - 1	PBS	159	0.0266	-7471B	PPB	03 Dec 2020 14:06:57	U	SMPL
PB133365BS - 1	LCSS	17956	3.3180	-7471B	PPB	03 Dec 2020 14:09:14	U	SMPL
L4915-01 - 1	OR-03-12220	1899	0.3484	-7471B	PPB	03 Dec 2020 14:11:30	U	SMPL
L4915-01DUP - 1	OR-03-12220DUP	1737	0.3184	-7471B	PPB	03 Dec 2020 14:13:46	U	SMPL
L4915-01MS - 1	OR-03-12220MS	19570	3.6165	-7471B	PPB	03 Dec 2020 14:16:03	U	SMPL
L4915-01MSD - 1	OR-03-12220MSD	20802	3.8443	-7471B	PPB	03 Dec 2020 14:18:20	U	SMPL
L4917-01 - 1	CL-02-12220	1857	0.3406	-7471B	PPB	03 Dec 2020 14:20:35	U	SMPL
CCV23 - 1	CCV23	26838	4.9606	-7471B	PPB	03 Dec 2020 14:22:51	U	SMPL
CCB23 - 1	CCB23	-62	-0.0143	-7471B	PPB	03 Dec 2020 14:25:06	U	SMPL
L4918-02 - 1	20801	1139	0.2079	-7471B	PPB	03 Dec 2020 14:27:21	U	SMPL
L4918-04 - 1	16827	665	0.1202	-7471B	PPB	03 Dec 2020 14:29:37	U	SMPL
L4918-06 - 1	16829	669	0.1209	-7471B	PPB	03 Dec 2020 14:31:53	U	SMPL
L4931-01 - 1	R-LA-33-1-WC	1123	0.2049	-7471B	PPB	03 Dec 2020 14:34:09	U	SMPL
L4932-01 - 1	R-LA-35-1-WC	2860	0.5261	-7471B	PPB	03 Dec 2020 14:36:25	U	SMPL
L4915-01LX5 - 1		813	0.1476	-7471B	PPB	03 Dec 2020 14:38:41	U	SMPL
L4915-01A - 1		22153	4.0942	-7471B	PPB	03 Dec 2020 14:40:57	U	SMPL
CCV24 - 1	CCV24	27114	5.0116	-7471B	PPB	03 Dec 2020 14:43:15	U	SMPL
CCB24 - 1	CCB24	-68	-0.0154	-7471B	PPB	03 Dec 2020 14:45:31	U	SMPL