



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112211

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	12/03/2020	16:05
2	CAL2	1	Water	NA	NA	21.2	7.00	12/03/2020	16:06
3	CAL3	1	Water	NA	NA	21.2	10.03	12/03/2020	16:07
4	ICV	1	Water	NA	NA	21.2	7.01	12/03/2020	16:08
5	CCV1	1	Water	NA	NA	21.2	2.01	12/03/2020	16:10
6	L4936-02	1	Solid	20.02	20	20.6	8.30	12/03/2020	16:11
7	L4936-02DUP	1	Solid	20.03	20	20.7	8.31	12/03/2020	16:12
8	CCV2	1	Water	NA	NA	21.2	12.03	12/03/2020	16:13

112211

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L4936 WorkList ID : 144655 Department : Wet-Chemistry Date : 12-03-2020 15:58:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/03/2020	Solid	L4936-02	Corrosivity	Cool 4 deg C	PSEG03	J33	VNJ-214	12/03/2020	9045D

Date/Time 12/03/2020 16:00
 Received by: J.R. (wet) Oreny
 Relinquished by: AP (Ext. Lab)

Date/Time 12.03.2020 17:00
 Received by: AP (Ext. Lab)
 Relinquished by: J.P. (Ext. Lab) (Oreny)