



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112215

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	12/04/2020	08:35
2	CAL2	1	Water	NA	NA	22.1	7.01	12/04/2020	08:36
3	CAL3	1	Water	NA	NA	22.1	10.04	12/04/2020	08:37
4	ICV	1	Water	NA	NA	22.2	7.00	12/04/2020	08:39
5	CCV1	1	Water	NA	NA	22.1	2.01	12/04/2020	08:40
6	L4593-01	1	Water	NA	NA	21.7	10.38	12/04/2020	08:42
7	L4593-01DUP	1	Water	NA	NA	21.8	10.39	12/04/2020	08:43
8	CCV2	1	Water	NA	NA	22.2	12.02	12/04/2020	08:45

112215

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14593 WorkList ID : 144666 Department : Wet-Chemistry Date : 12-04-2020 08:27:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/03/2020	Water	L4593-01- 1	pH	Cool 4 deg C	TULL02	J51	BARNUM-STREET-TP-1	12/03/2020	9040C

Date/Time 12-04-2020 8:30
Received by: JB (604) Chem
Relinquished by: CP (SM)

Date/Time 12-04-2020 9:45
Received by: CP (SM)
Relinquished by: JB (604) Chem