



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112219

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	12/04/2020	11:02
2	L4593-01		1	8	100.00	>212.0	>212.0	12/04/2020	11:35
3	L4593-01DUP		1	8	100.00	>212.0	>212.0	12/04/2020	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB112219

WorkList Name : fp 12420 WorkList ID : 144680 Department : Wet-Chemistry Date : 12-04-2020 09:11:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2020	Water	L4593-01 A	Flash Point	Cool 4 deg C	TULL02	J51	BARNUM-STREET-TP-1	12/03/2020	1010A

Date/Time 12/4/20 10:30
 Received by: KS (Webster)
 Relinquished by: RJ (Ext-1ab)

Date/Time 12/4/20 12:15
 Received by: RJ (Ext-1ab)
 Relinquished by: KS (Webster)