

LB11222


INSTRUMENT ID: CV1

245.1

Linear

 μ Abs.:

32307



A= 0.0000e+000

B= 3.0699e-004

C= -5.6508e-002

Rho= 0.9997321

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	140	0.000	140				
0.05	0.050	0.046	-0.004	333	0.0 %	333	0			
0.2	0.200	0.181	-0.019	774	0.0 %	774				
2.5	2.500	2.438	-0.062	8127	0.0 %	8127				
5.0	5.000	5.114	0.114	16842	0.0 %	16842				
7.5	7.500	7.623	0.123	25016	0.0 %	25016				
10.0	10.000	9.861	-0.139	32307	0.0 %	32307				

LB112222 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 04 Dec 2020 12:01:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	140	-	0.0000	245.1	PPB	04 Dec 2020 12:16:23	S	Std
0.05 - 1	5005	333	-	0.0500	245.1	PPB	04 Dec 2020 12:18:40	S	Std
0.2 - 1	502	774	-	0.2000	245.1	PPB	04 Dec 2020 12:20:57	S	Std
2.5 - 1	52.5	8127	-	2.5000	245.1	PPB	04 Dec 2020 12:23:13	S	Std
5.0 - 1	55	16842	-	5.0000	245.1	PPB	04 Dec 2020 12:25:30	S	Std
7.5 - 1	57.5	25016	-	7.5000	245.1	PPB	04 Dec 2020 12:27:46	S	Std
10.0 - 1	516	32307	-	10.0000	245.1	PPB	04 Dec 2020 12:34:00	S	Std
ICV27 - 1	ICV27	13844	4.1934	-245.1		PPB	04 Dec 2020 12:39:16	U	SMPL
ICB27 - 1	ICB27	109	-0.0230	-245.1		PPB	04 Dec 2020 12:41:31	U	SMPL
CCV25 - 1	CCV25	16870	5.1224	-245.1		PPB	04 Dec 2020 12:43:47	U	SMPL
CCB25 - 1	CCB25	108	-0.0234	-245.1		PPB	04 Dec 2020 12:46:02	U	SMPL
CRA - 1	CRA	320	0.0417	-245.1		PPB	04 Dec 2020 12:48:18	U	SMPL
HighStd - 1	HighStd	32347	9.8737	-245.1		PPB	04 Dec 2020 12:50:33	U	SMPL
ChkStd - 1	ChkStd	24224	7.3800	-245.1		PPB	04 Dec 2020 12:52:49	U	SMPL
PB133404BL - 1	PBW	138	-0.0141	-245.1		PPB	04 Dec 2020 12:57:42	U	SMPL
PB133404BS - 1	LCSW	13962	4.2297	-245.1		PPB	04 Dec 2020 12:59:57	U	SMPL
L4919-02 - 1	COMP	82	-0.0313	-245.1		PPB	04 Dec 2020 13:05:03	U	SMPL
L4919-02DUP - 1	COMPDU	72	-0.0344	-245.1		PPB	04 Dec 2020 13:07:20	U	SMPL
L4919-02MS - 1	COMPMS	12600	3.8116	-245.1		PPB	04 Dec 2020 13:09:36	U	SMPL
L4919-02MSD - 1	COMPMSD	12841	3.8855	-245.1		PPB	04 Dec 2020 13:11:54	U	SMPL
L4938-01 - 1	IB1-TREATED-EFFLU	288	0.0319	-245.1		PPB	04 Dec 2020 13:14:09	U	SMPL
CCV26 - 1	CCV26	16882	5.1261	-245.1		PPB	04 Dec 2020 13:16:24	U	SMPL
CCB26 - 1	CCB26	108	-0.0234	-245.1		PPB	04 Dec 2020 13:18:40	U	SMPL
L4938-02 - 1	IB1-TREATED-EFFLU	219	0.0107	-245.1		PPB	04 Dec 2020 13:20:56	U	SMPL
L4919-02LX5 - 1		169	-0.0046	-245.1		PPB	04 Dec 2020 13:23:11	U	SMPL
L4919-02A - 1		13056	3.9515	-245.1		PPB	04 Dec 2020 13:25:27	U	SMPL
CCV27 - 1	CCV27	16882	5.1261	-245.1		PPB	04 Dec 2020 13:27:42	U	SMPL
CCB27 - 1	CCB27	87	-0.0298	-245.1		PPB	04 Dec 2020 13:29:58	U	SMPL