



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112223

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	12/04/2020	12:10
2	CAL2	1	Water	NA	NA	21.3	7.01	12/04/2020	12:11
3	CAL3	1	Water	NA	NA	21.3	10.03	12/04/2020	12:12
4	ICV	1	Water	NA	NA	21.3	6.99	12/04/2020	12:15
5	CCV1	1	Water	NA	NA	21.3	2.01	12/04/2020	12:16
6	L4598-01	1	Water	NA	NA	19.9	5.74	12/04/2020	12:20
7	L4598-01DUP	1	Water	NA	NA	19.8	5.75	12/04/2020	12:21
8	L4598-02	1	Water	NA	NA	19.9	5.36	12/04/2020	12:22
9	L4598-03	1	Water	NA	NA	19.4	5.41	12/04/2020	12:25
10	CCV2	1	Water	NA	NA	21.3	12.02	12/04/2020	12:30

2B 112223

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4598 WorkList ID : 144684 Department : Wet-Chemistry Date : 12-04-2020 11:53:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/03/2020	Water	L4598-01	pH	Cool 4 deg C	STER07	J61	CELL-A	12/03/2020	9040C
12/03/2020	Water	L4598-02	pH	Cool 4 deg C	STER07	J61	CELLS-B/C	12/03/2020	9040C
12/03/2020	Water	L4598-03	pH	Cool 4 deg C	STER07	J61	CELL-D	12/03/2020	9040C

Date/Time 12/04/2020 12:11:00
Received by: CP (SMI)
Relinquished by: CP (SMI)

Date/Time 12/04/2020 13:10:06
Received by: CP (SMI)
Relinquished by: CP (SMI)