



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112236

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	12/04/2020	15:24
2	CAL2	1	Water	NA	NA	21.2	7.01	12/04/2020	15:25
3	CAL3	1	Water	NA	NA	21.2	10.02	12/04/2020	15:26
4	ICV	1	Water	NA	NA	21.2	6.98	12/04/2020	15:30
5	CCV1	1	Water	NA	NA	21.2	2.01	12/04/2020	15:32
6	L4962-01	1	Water	NA	NA	15.1	6.95	12/04/2020	15:34
7	L4962-01DUP	1	Water	NA	NA	15.2	6.96	12/04/2020	15:35
8	L4962-02	1	Water	NA	NA	15.3	6.96	12/04/2020	15:37
9	CCV2	1	Water	NA	NA	21.3	12.03	12/04/2020	15:39

VB 112236

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4962 WorkList ID : 144697 Department : Wet-Chemistry Date : 12-04-2020 15:21:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/03/2020	Water	L4962-01	-F pH	Cool 4 deg C	LOCK01	J61	MW-2	12/03/2020	SM 4500-H B
12/03/2020	Water	L4962-02	-F pH	Cool 4 deg C	LOCK01	J61	MW-4	12/03/2020	SM 4500-H B

Date/Time 12-04-2020 15:20
 Received by: JO (Infpm)
 Relinquished by: JO (Sul)

Date/Time 12-04-2020 16:30
 Received by: JO (Sul)
 Relinquished by: JO (Sul) *Cherny*