



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB112239

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP88442
Turbidity Calibration std, 20NTU	WP88438
Turbidity Calibration std, 40NTU	WP88437
Turbidity Calibration std, 1NTU	WP88441
Turbidity Calibration std, 5NTU	WP88440
Turbidity Calibration - CCV std, 10 NTU	WP88439

Intercept: 0.234

Slope: 0.9604

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.262	12/04/2020	15:30
2	CAL2	1	1	1.180	12/04/2020	15:32
3	CAL3	5	1	5.230	12/04/2020	15:34
4	CAL4	10	1	9.820	12/04/2020	15:36
5	CAL5	20	1	19.100	12/04/2020	15:38
6	CAL6	40	1	38.800	12/04/2020	15:40



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB112239

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.300	10.481	12/04/2020	15:43
2	ICB		1	0.280	0.048	12/04/2020	15:45
3	CCV1	10	1	9.540	9.690	12/04/2020	15:48
4	CCB1		1	0.285	0.053	12/04/2020	15:50
5	LB112239BL		1	0.274	0.042	12/04/2020	15:52
6	L4962-01		1	13.800	14.125	12/04/2020	15:55
7	L4962-01DUP		1	14.100	14.438	12/04/2020	15:57
8	L4962-02		1	15.100	15.479	12/04/2020	15:59
9	CCV2	10	1	9.920	10.085	12/04/2020	16:02
10	CCB2		1	0.297	0.066	12/04/2020	16:04

LB112239
 WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-120420 WorkList ID : 144695 Department : Wet-Chemistry Date : 12-04-2020 14:35:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2020	Water	L4962-01	Turbidity	Cool 4 deg C	LOCK01	J61	MW-2	12/03/2020	SM2130 B
12/05/2020	Water	L4962-02	Turbidity	Cool 4 deg C	LOCK01	J61	MW-4	12/03/2020	SM2130 B

Date/Time
 Received by:
 Relinquished by:

12/04/2020 14:50
 AMC wetchem
 7D (29-2000)

Date/Time
 Received by:
 Relinquished by:

12/04/2020 16:10
 7D (29-2000)
 AMC wetchem