

LB112241


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49462



A= 0.0000e+000

B= 2.0299e-004

C= -2.1284e-002

Rho= 0.9998988

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	113	0.000	113				
0.2	0.200	0.194	-0.006	1063	0.0 %	1063				
2.5	2.500	2.471	-0.029	12280	0.0 %	12280				
5.0	5.000	5.094	0.094	25199	0.0 %	25199				
7.5	7.500	7.419	-0.081	36655	0.0 %	36655				
10.0	10.000	10.019	0.019	49462	0.0 %	49462				

LB112241 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Dec 2020 16:38:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	113	-	0.0000	7471B	PPB	04 Dec 2020 16:39:15	S	Std
0.2 - 1	50.2	1063	-	0.2000	7471B	PPB	04 Dec 2020 16:41:32	S	Std
2.5 - 1	52.5	12280	-	2.5000	7471B	PPB	04 Dec 2020 16:43:49	S	Std
5.0 - 1	55	25199	-	5.0000	7471B	PPB	04 Dec 2020 16:46:07	S	Std
7.5 - 1	57.5	36655	-	7.5000	7471B	PPB	04 Dec 2020 16:48:25	S	Std
10.0 - 1	510	49462	-	10.0000	7471B	PPB	04 Dec 2020 16:50:41	S	Std
ICV29 - 1	ICV29	20984	3.8780	-	7471B	PPB	04 Dec 2020 16:56:03	U	SMPL
ICB29 - 1	ICB29	55	0.0074	-	7471B	PPB	04 Dec 2020 16:58:18	U	SMPL
CCV31 - 1	CCV31	25247	5.1037	-	7471B	PPB	04 Dec 2020 17:00:55	U	SMPL
CCB31 - 1	CCB31	61	-0.0089	-	7471B	PPB	04 Dec 2020 17:03:10	U	SMPL
CRA - 1	CRA	1033	0.1884	-	7471B	PPB	04 Dec 2020 17:05:25	U	SMPL
HighStd - 1	HighStd	50007	10.1297	-	7471B	PPB	04 Dec 2020 17:07:40	U	SMPL
ChkStd - 1	ChkStd	34738	7.0302	-	7471B	PPB	04 Dec 2020 17:09:56	U	SMPL
PB133406BL - 1	PBS	111	0.0012	-	7471B	PPB	04 Dec 2020 17:12:13	U	SMPL
PB133406BS - 1	LCS	18756	3.7860	-	7471B	PPB	04 Dec 2020 17:14:30	U	SMPL
L4845-09 - 1	PH-TP	40266	8.1524	-	7471B	PPB	04 Dec 2020 17:16:46	U	SMPL
L4845-09DUP - 1	PH-TPDUP	47453	9.6113	-	7471B	PPB	04 Dec 2020 17:19:03	U	SMPL
L4845-09MS - 1	PH-TPMS	68061	13.7945	-	7471B	PPB	04 Dec 2020 17:26:28	U	SMPL
L4845-09MSD - 1	PH-TPMSD	61492	12.4611	-	7471B	PPB	04 Dec 2020 17:28:46	U	SMPL
L4916-01 - 1	OK-01-12220	1254	0.2333	-	7471B	PPB	04 Dec 2020 17:31:04	U	SMPL
CCV32 - 1	CCV32	25295	5.1134	-	7471B	PPB	04 Dec 2020 17:33:21	U	SMPL
CCB32 - 1	CCB32	71	-0.0069	-	7471B	PPB	04 Dec 2020 17:35:37	U	SMPL
L4935-01 - 1	OK-02-12220	1313	0.2452	-	7471B	PPB	04 Dec 2020 17:37:52	U	SMPL
L4936-01 - 1	VNJ-214	3932	0.7769	-	7471B	PPB	04 Dec 2020 17:40:07	U	SMPL
L4939-01 - 1	EO-02-12320	754	0.1318	-	7471B	PPB	04 Dec 2020 17:42:23	U	SMPL
L4845-09LX5 - 1		7993	1.6012	-	7471B	PPB	04 Dec 2020 17:44:39	U	SMPL
L4845-09A - 1		60263	12.2116	-	7471B	PPB	04 Dec 2020 17:46:54	U	SMPL
CCV33 - 1	CCV33	25351	5.1248	-	7471B	PPB	04 Dec 2020 17:49:11	U	SMPL
CCB33 - 1	CCB33	57	-0.0097	-	7471B	PPB	04 Dec 2020 17:51:29	U	SMPL
L4845-09DLX2 - 1		20418	4.1234	-	7471B	PPB	04 Dec 2020 17:53:47	U	SMPL
L4845-09DUPDLX2 - 1		24025	4.8556	-	7471B	PPB	04 Dec 2020 17:56:03	U	SMPL
L4845-09MSDLX2 - 1		33111	6.7000	-	7471B	PPB	04 Dec 2020 17:58:20	U	SMPL
L4845-09MSDDLX2 - 1		31392	6.3510	-	7471B	PPB	04 Dec 2020 18:00:35	U	SMPL
L4845-09LDLX10 - 1		3965	0.7836	-	7471B	PPB	04 Dec 2020 18:02:52	U	SMPL
L4845-0ADLX2 - 1		30937	6.2587	-	7471B	PPB	04 Dec 2020 18:05:09	U	SMPL
CCV34 - 1	CCV34	25232	5.1006	-	7471B	PPB	04 Dec 2020 18:07:25	U	SMPL
CCB34 - 1	CCB34	15	-0.0182	-	7471B	PPB	04 Dec 2020 18:09:41	U	SMPL