



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112249

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	12/07/2020	12:45
2	CAL2	1	Water	NA	NA	21.3	7.00	12/07/2020	12:46
3	CAL3	1	Water	NA	NA	21.3	10.03	12/07/2020	12:47
4	ICV	1	Water	NA	NA	21.3	6.99	12/07/2020	12:48
5	CCV1	1	Water	NA	NA	21.3	2.01	12/07/2020	12:49
6	L4962-05	1	Water	NA	NA	16.0	6.14	12/07/2020	12:50
7	L4962-05DUP	1	Water	NA	NA	16.1	6.15	12/07/2020	12:51
8	L4962-06	1	Water	NA	NA	16.9	6.85	12/07/2020	12:53
9	CCV2	1	Water	NA	NA	21.2	12.02	12/07/2020	13:00

WORKLIST(Hardcopy Internal Chain)

112249

WorkList Name : PH L4962-2 WorkList ID : 144729 Department : Wet-Chemistry Date : 12-07-2020 12:36:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/03/2020	Water	L4962-05	pH	Cool 4 deg C	LOCK01	J61	MW-1	12/03/2020	SM 4500-H B
12/03/2020	Water	L4962-06	pH	Cool 4 deg C	LOCK01	J61	MW-3	12/03/2020	SM 4500-H B

Date/Time 12/07/2020 12:40
 Received by: JP/Wet Chem
 Relinquished by: CP SM1

Date/Time 12/07/2020 12:40
 Received by: CP SM
 Relinquished by: JP/Wet Chem