



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB112253

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP88465
Turbidity Calibration std, 20NTU	WP88461
Turbidity Calibration std, 40NTU	WP88460
Turbidity Calibration std, 1NTU	WP88464
Turbidity Calibration std, 5NTU	WP88463
Turbidity Calibration - CCV std, 10 NTU	WP88462
10 NTU Standard 500 ml	W2752

Intercept: 0.142

Slope: 0.9711

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.267	12/07/2020	13:20
2	CAL2	1	1	1.090	12/07/2020	13:22
3	CAL3	5	1	5.340	12/07/2020	13:24
4	CAL4	10	1	9.560	12/07/2020	13:26
5	CAL5	20	1	19.200	12/07/2020	13:28
6	CAL6	40	1	39.200	12/07/2020	13:30



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB112253

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	10.254	12/07/2020	13:33
2	ICB		1	0.282	0.144	12/07/2020	13:35
3	CCV1	10	1	9.770	9.915	12/07/2020	13:37
4	CCB1		1	0.288	0.150	12/07/2020	13:39
5	LB112253BL		1	0.302	0.165	12/07/2020	13:42
6	L4962-05		1	0.567	0.438	12/07/2020	13:44
7	L4962-06		1	1.320	1.213	12/07/2020	13:46
8	L4962-06DUP		1	1.430	1.326	12/07/2020	13:48
9	CCV2	10	1	9.830	9.976	12/07/2020	13:51
10	CCB2		1	0.292	0.154	12/07/2020	13:53

LB112283
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-120720 WorkList ID : 144728 Department : Wet-Chemistry Date : 12-07-2020 12:19:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2020	Water	L4962-05 F	Turbidity	Cool 4 deg C	LOCK01	J61	MW-1	12/03/2020	SM2130 B
12/05/2020	Water	L4962-06 P	Turbidity	Cool 4 deg C	LOCK01	J61	MW-3	12/03/2020	SM2130 B

Date/Time 12/07/2020 12:58
 Received by: Am (Wet-Chem)
 Relinquished by: SB/coat (chem)

Date/Time 12/07/2020 13:55
 Received by: SB/coat (chem)
 Relinquished by: Am (Wet-Chem)