



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112273

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.00	12/08/2020	14:05
2	CAL2	1	Water	NA	NA	21.2	7.01	12/08/2020	14:06
3	CAL3	1	Water	NA	NA	21.2	10.03	12/08/2020	14:09
4	ICV	1	Water	NA	NA	21.2	7.00	12/08/2020	14:10
5	CCV1	1	Water	NA	NA	21.2	2.02	12/08/2020	14:11
6	L4975-01	1	Solid	20.03	20	21.3	5.52	12/08/2020	14:15
7	L4975-01DUP	1	Solid	20.04	20	21.2	5.53	12/08/2020	14:16
8	CCV2	1	Water	NA	NA	21.3	12.02	12/08/2020	14:18

WORKLIST(Hardcopy Internal Chain)

HB 112273

WorkList Name : PH L4975 WorkList ID : 144758 Department : Wet-Chemistry Date : 12-08-2020 13:51:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/08/2020	Solid	L4975-01 -C	pH	Cool 4 deg C	PSEG03	L13	21222	12/08/2020	9045D

Date/Time 12/08/2020 14:00
 Received by: JOLLY CHAM
 Relinquished by: AT SM

Date/Time 12-08-2020 15:00
 Received by: AT SM
 Relinquished by: JOLLY CHAM