


LB112287

7471B

INSTRUMENT ID: CV2

 μ Abs.:

53488



A= 0.0000e+000

B= 1.8433e-004

C= -9.5276e-003

Rho= 0.9995793

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	62	0.000	62				
0.2	0.200	0.196	-0.004	1115	0.0 %	1115				
2.5	2.500	2.397	-0.103	13057	0.0 %	13057				
5.0	5.000	5.089	0.089	27659	0.0 %	27659				
7.5	7.500	7.666	0.166	41643	0.0 %	41643				
10.0	10.000	9.850	-0.150	53488	0.0 %	53488				

LB112287 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Dec 2020 15:49:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	62	-	0.0000	7471B	PPB	09 Dec 2020 15:50:49	S	Std
0.2 - 1	50.2	1115	-	0.2000	7471B	PPB	09 Dec 2020 15:53:06	S	Std
2.5 - 1	52.5	13057	-	2.5000	7471B	PPB	09 Dec 2020 15:55:24	S	Std
5.0 - 1	55	27659	-	5.0000	7471B	PPB	09 Dec 2020 15:57:41	S	Std
7.5 - 1	57.5	41643	-	7.5000	7471B	PPB	09 Dec 2020 15:59:57	S	Std
10.0 - 1	510	53488	-	10.0000	7471B	PPB	09 Dec 2020 16:04:50	S	Std
ICV34 - 1	ICV34	21041	3.8689	-7471B	PPB	PPB	09 Dec 2020 16:07:33	U	SMPL
ICB34 - 1	ICB34	-43	-0.0175	-7471B	PPB	PPB	09 Dec 2020 16:09:47	U	SMPL
CCV49 - 1	CCV49	27516	5.0624	-7471B	PPB	PPB	09 Dec 2020 16:12:02	U	SMPL
CCB49 - 1	CCB49	-52	-0.0191	-7471B	PPB	PPB	09 Dec 2020 16:14:17	U	SMPL
CRA - 1	CRA	1090	0.1914	-7471B	PPB	PPB	09 Dec 2020 16:16:33	U	SMPL
HighStd - 1	HighStd	54207	9.9823	-7471B	PPB	PPB	09 Dec 2020 16:18:48	U	SMPL
ChkStd - 1	ChkStd	38792	7.1409	-7471B	PPB	PPB	09 Dec 2020 16:21:03	U	SMPL
PB133498BL - 1	PBS	-108	-0.0294	-7471B	PPB	PPB	09 Dec 2020 16:23:20	U	SMPL
PB133498BS - 1	LCSS	19313	3.5504	-7471B	PPB	PPB	09 Dec 2020 16:25:37	U	SMPL
L4972-01 - 1	SU-01-120820	5433	0.9919	-7471B	PPB	PPB	09 Dec 2020 16:27:54	U	SMPL
L4972-01DUP - 1	SU-01-120820DUP	5235	0.9554	-7471B	PPB	PPB	09 Dec 2020 16:30:10	U	SMPL
L4972-01MS - 1	SU-01-120820MS	24530	4.5120	-7471B	PPB	PPB	09 Dec 2020 16:32:27	U	SMPL
L4972-01MSD - 1	SU-01-120820MSD	27659	5.0888	-7471B	PPB	PPB	09 Dec 2020 16:34:44	U	SMPL
L4973-01 - 1	HD-01-120820	3949	0.7184	-7471B	PPB	PPB	09 Dec 2020 16:37:00	U	SMPL
CCV50 - 1	CCV50	28143	5.1780	-7471B	PPB	PPB	09 Dec 2020 16:39:15	U	SMPL
CCB50 - 1	CCB50	-37	-0.0163	-7471B	PPB	PPB	09 Dec 2020 16:41:31	U	SMPL
L4977-01 - 1	72-11976	3977	0.7235	-7471B	PPB	PPB	09 Dec 2020 16:43:47	U	SMPL
L4982-01 - 1	R-LA-36-5-WC	2463	0.4445	-7471B	PPB	PPB	09 Dec 2020 16:46:02	U	SMPL
L4983-01 - 1	TR1-A-D-COMP	1895	0.3398	-7471B	PPB	PPB	09 Dec 2020 16:48:18	U	SMPL
L4983-12 - 1	TR2-A-D-COMP	1217	0.2148	-7471B	PPB	PPB	09 Dec 2020 16:50:33	U	SMPL
L4984-01 - 1	TR3-A-D-COMP	1201	0.2118	-7471B	PPB	PPB	09 Dec 2020 16:52:49	U	SMPL
L4972-01LX5 - 1		909	0.1580	-7471B	PPB	PPB	09 Dec 2020 16:55:05	U	SMPL
L4972-01A - 1		26713	4.9144	-7471B	PPB	PPB	09 Dec 2020 16:57:22	U	SMPL
CCV51 - 1	CCV51	27789	5.1127	-7471B	PPB	PPB	09 Dec 2020 16:59:39	U	SMPL
CCB51 - 1	CCB51	-34	-0.0158	-7471B	PPB	PPB	09 Dec 2020 17:01:57	U	SMPL