



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112290

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	12/09/2020	17:10
2	CAL2	1	Water	NA	NA	21.2	7.00	12/09/2020	17:11
3	CAL3	1	Water	NA	NA	21.2	10.03	12/09/2020	17:12
4	ICV	1	Water	NA	NA	21.2	6.99	12/09/2020	17:15
5	CCV1	1	Water	NA	NA	21.3	2.01	12/09/2020	17:16
6	L4985-05	1	Water	NA	NA	10.7	5.50	12/09/2020	17:18
7	L4985-05DUP	1	Water	NA	NA	10.8	5.51	12/09/2020	17:19
8	L4986-03	1	Water	NA	NA	22.9	6.87	12/09/2020	17:20
9	L4987-01	1	Water	NA	NA	12.9	8.70	12/09/2020	17:21
10	L4987-03	1	Water	NA	NA	11.0	5.46	12/09/2020	17:22
11	L4987-05	1	Water	NA	NA	11.8	5.63	12/09/2020	17:25
12	L4988-03	1	Water	NA	NA	22.0	6.02	12/09/2020	17:30
13	L4991-01	1	Water	NA	NA	13.7	8.66	12/09/2020	17:31
14	L4991-03	1	Water	NA	NA	15.0	5.72	12/09/2020	17:32
15	L4991-04	1	Water	NA	NA	15.7	5.50	12/09/2020	17:35
16	CCV2	1	Water	NA	NA	21.3	12.02	12/09/2020	17:36
17	L4991-05	1	Water	NA	NA	17.1	6.58	12/09/2020	17:29
18	L4991-06	1	Water	NA	NA	16.5	5.89	12/09/2020	17:30
19	L4991-07	1	Water	NA	NA	14.1	8.32	12/09/2020	17:32
20	L4991-09	1	Water	NA	NA	16.8	7.72	12/09/2020	17:35

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	L4991-10	1	Water	NA	NA	16.2	8.40	12/09/2020	17:36
22	L4991-11	1	Water	NA	NA	16.6	8.69	12/09/2020	17:29
23	L4992-01	1	Water	NA	NA	13.9	6.79	12/09/2020	17:30
24	CCV3	1	Water	NA	NA	21.3	2.01	12/09/2020	17:32

28 112296

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14992 WorkList ID : 144785 Department : Wet-Chemistry Date : 12-09-2020 16:50:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/09/2020	Water	L4985-05	pH	Cool 4 deg C	PSEG03	L21	32831	12/09/2020	9040C
12/09/2020	Water	L4986-03	pH	Cool 4 deg C	PSEG03	L21	32794	12/09/2020	9040C
12/09/2020	Water	L4987-01	pH	Cool 4 deg C	PSEG03	L21	32833	12/09/2020	9040C
12/09/2020	Water	L4987-03	pH	Cool 4 deg C	PSEG03	L21	32832	12/09/2020	9040C
12/09/2020	Water	L4987-05	pH	Cool 4 deg C	PSEG03	L21	32834	12/09/2020	9040C
12/09/2020	Water	L4988-03	pH	Cool 4 deg C	PSEG03	L22	32818	12/09/2020	9040C
12/09/2020	Water	L4991-01	pH	Cool 4 deg C	PSEG03	L21	COMP-2	12/09/2020	9040C
12/09/2020	Water	L4991-03	pH	Cool 4 deg C	PSEG03	L21	31340	12/09/2020	9040C
12/09/2020	Water	L4991-04	pH	Cool 4 deg C	PSEG03	L21	31341	12/09/2020	9040C
12/09/2020	Water	L4991-05	pH	Cool 4 deg C	PSEG03	L21	31344	12/09/2020	9040C
12/09/2020	Water	L4991-06	pH	Cool 4 deg C	PSEG03	L21	31339	12/09/2020	9040C
12/09/2020	Water	L4991-07	pH	Cool 4 deg C	PSEG03	L21	COMP-1	12/09/2020	9040C
12/09/2020	Water	L4991-09	pH	Cool 4 deg C	PSEG03	L21	31345	12/09/2020	9040C
12/09/2020	Water	L4991-10	pH	Cool 4 deg C	PSEG03	L21	31346	12/09/2020	9040C
12/09/2020	Water	L4991-11	pH	Cool 4 deg C	PSEG03	L21	31347	12/09/2020	9040C
12/09/2020	Water	L4992-01	pH	Cool 4 deg C	PSEG03	L22	TRE-1130	12/09/2020	9040C

Date/Time 12.09.2020 17:00
 Received by: JH (wet-chem)
 Relinquished by: KS (wet-chem)

Date/Time 12.09.2020 18:25
 Received by: KS (wet-chem)
 Relinquished by: JH (wet-chem)