


LB112298

INSTRUMENT ID: CV1

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-21	0.000	-21				
0.2	0.200	0.181	-0.019	679	0.0 %	679				
2.5	2.500	2.439	-0.061	8971	0.0 %	8971				
5.0	5.000	5.092	0.092	18712	0.0 %	18712				
7.5	7.500	7.633	0.133	28042	0.0 %	28042				
10.0	10.000	9.864	-0.136	36233	0.0 %	36233				

LB112298 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Dec 2020 11:43:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.0000	7470A	PPB	10 Dec 2020 11:44:43	S	Std
0.2 - 1	50.2	679	-	0.2000	7470A	PPB	10 Dec 2020 11:47:00	S	Std
2.5 - 1	52.5	8971	-	2.5000	7470A	PPB	10 Dec 2020 11:49:18	S	Std
5.0 - 1	55	18712	-	5.0000	7470A	PPB	10 Dec 2020 11:51:35	S	Std
7.5 - 1	57.5	28042	-	7.5000	7470A	PPB	10 Dec 2020 11:53:51	S	Std
10.0 - 1	61.0	36233	-	10.0000	7470A	PPB	10 Dec 2020 11:56:07	S	Std
ICV35 - 1	ICV35	13712	3.7304	-	7470A	PPB	10 Dec 2020 11:58:51	U	SMPL
ICB35 - 1	ICB35	-65	-0.0220	-	7470A	PPB	10 Dec 2020 12:01:06	U	SMPL
CCV52 - 1	CCV52	18562	5.0514	-	7470A	PPB	10 Dec 2020 12:03:21	U	SMPL
CCB52 - 1	CCB52	-62	-0.0212	-	7470A	PPB	10 Dec 2020 12:05:36	U	SMPL
CRA - 1	CRA	723	0.1926	-	7470A	PPB	10 Dec 2020 12:07:51	U	SMPL
HighStd - 1	HighStd	36365	9.9004	-	7470A	PPB	10 Dec 2020 12:10:06	U	SMPL
ChkStd - 1	ChkStd	25488	6.9379	-	7470A	PPB	10 Dec 2020 12:12:22	U	SMPL
PB133522BL - 1	PBW	-84	-0.0272	-	7470A	PPB	10 Dec 2020 12:14:38	U	SMPL
PB133522BS - 1	LCSW	13887	3.7781	-	7470A	PPB	10 Dec 2020 12:16:54	U	SMPL
L4913-02 - 1	TEST-PIT-3	-50	-0.0179	-	7470A	PPB	10 Dec 2020 12:19:11	U	SMPL
L4975-01 - 1	21222	33	0.0047	-	7470A	PPB	10 Dec 2020 12:21:27	U	SMPL
L4976-01 - 1	DRUM-1	-7	-0.0062	-	7470A	PPB	10 Dec 2020 12:23:44	U	SMPL
L4976-02 - 1	DRUM-2	42	0.0071	-	7470A	PPB	10 Dec 2020 12:26:01	U	SMPL
L4976-03 - 1	DRUM-3	20	0.0012	-	7470A	PPB	10 Dec 2020 12:28:16	U	SMPL
CCV53 - 1	CCV53	18197	4.9520	-	7470A	PPB	10 Dec 2020 12:30:31	U	SMPL
CCB53 - 1	CCB53	-50	-0.0179	-	7470A	PPB	10 Dec 2020 12:32:46	U	SMPL
L4983-02 - 1	TR1-A-D-COMP	9	-0.0018	-	7470A	PPB	10 Dec 2020 12:35:02	U	SMPL
L4983-13 - 1	TR2-A-D-COMP	23	0.0020	-	7470A	PPB	10 Dec 2020 12:37:17	U	SMPL
L4984-02 - 1	TR3-A-D-COMP	12	-0.0010	-	7470A	PPB	10 Dec 2020 12:39:32	U	SMPL
L4984-02DUP - 1	TR3-A-D-COMPDUP	30	0.0039	-	7470A	PPB	10 Dec 2020 12:41:48	U	SMPL
L4984-02MS - 1	TR3-A-D-COMPMS	14298	3.8900	-	7470A	PPB	10 Dec 2020 12:44:03	U	SMPL
L4984-02MSD - 1	TR3-A-D-COMPMSD	13945	3.7939	-	7470A	PPB	10 Dec 2020 12:46:20	U	SMPL
PB133522TB - 1	PB133490TB	-37	-0.0144	-	7470A	PPB	10 Dec 2020 12:48:36	U	SMPL
L4984-02LX5 - 1		22	0.0017	-	7470A	PPB	10 Dec 2020 12:50:53	U	SMPL
L4984-02A - 1		15028	4.0889	-	7470A	PPB	10 Dec 2020 12:53:10	U	SMPL
CCV54 - 1	CCV54	18436	5.0171	-	7470A	PPB	10 Dec 2020 12:55:26	U	SMPL
CCB54 - 1	CCB54	-67	-0.0225	-	7470A	PPB	10 Dec 2020 12:57:41	U	SMPL